

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022324\
 Data File : VX040428.D
 Acq On : 23 Feb 2024 13:51
 Operator : JC/MD
 Sample : P1539-04RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20241105-AMENDED-COMPOSITE-35N-N-4-04RE

Quant Time: Feb 23 22:18:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022224W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 23 06:47:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	68552	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	121386	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	115309	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	58948	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	55011	49.274	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	98.540%		
35) Dibromofluoromethane	5.385	113	26893	32.877	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	65.760%#		
50) Toluene-d8	8.647	98	149951	50.750	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	101.500%		
62) 4-Bromofluorobenzene	11.079	95	57716	48.747	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	97.500%		
Target Compounds						
16) Acetone	2.380	43	26014	55.922	ug/l	97
20) Methylene Chloride	2.788	84	2845	2.921	ug/l #	80
25) 2-Butanone	4.568	43	5030	6.936	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022324\
Data File : VX040428.D
Acq On : 23 Feb 2024 13:51
Operator : JC/MD
Sample : P1539-04RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20241105-AMENDED-COMPOSITE-35N-N-4-04RE

Quant Time: Feb 23 22:18:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022224W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 23 06:47:29 2024
Response via : Initial Calibration

