

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022324\
 Data File : VX040431.D
 Acq On : 23 Feb 2024 15:00
 Operator : JC/MD
 Sample : PB159217ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB159217ZHE#01

Quant Time: Feb 23 22:19:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022224W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 23 06:47:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	74832	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	133439	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	134698	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	65922	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	60798	49.887	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.780%
35) Dibromofluoromethane	5.391	113	44275	49.238	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.480%
50) Toluene-d8	8.647	98	169347	52.138	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.280%
62) 4-Bromofluorobenzene	11.079	95	66996	51.474	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.940%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	21391	42.125	ug/l	95
20) Methylene Chloride	2.788	84	5331	5.014	ug/l	93
37) Ethyl Acetate	4.721	43	2455	1.544	ug/l #	86
43) Isopropyl Acetate	6.336	43	78022	31.981	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022324\
Data File : VX040431.D
Acq On : 23 Feb 2024 15:00
Operator : JC/MD
Sample : PB159217ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB159217ZHE#01

Quant Time: Feb 23 22:19:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022224W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 23 06:47:29 2024
Response via : Initial Calibration

