

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022324\
 Data File : VX040436.D
 Acq On : 23 Feb 2024 16:55
 Operator : JC/MD
 Sample : PB159217ZHE#06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB159217ZHE#06

Quant Time: Feb 23 22:21:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022224W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 23 06:47:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	73604	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	131184	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	131459	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	65803	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	59674	49.782	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 99.560%			
35) Dibromofluoromethane	5.385	113	44414	50.241	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.480%			
50) Toluene-d8	8.647	98	166543	52.156	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 104.320%			
62) 4-Bromofluorobenzene	11.079	95	66650	52.089	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 104.180%			
Target Compounds					Qvalue	
16) Acetone	2.379	43	24304	48.660	ug/l	95
20) Methylene Chloride	2.788	84	9600	9.180	ug/l	97
43) Isopropyl Acetate	6.342	43	54246	22.618	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022324\
Data File : VX040436.D
Acq On : 23 Feb 2024 16:55
Operator : JC/MD
Sample : PB159217ZHE#06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB159217ZHE#06

Quant Time: Feb 23 22:21:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022224W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 23 06:47:29 2024
Response via : Initial Calibration

