

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022324\
 Data File : VX040439.D
 Acq On : 23 Feb 2024 18:04
 Operator : JC/MD
 Sample : PB159217ZHE#09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB159217ZHE#09

Quant Time: Feb 23 22:23:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022224W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 23 06:47:29 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	66810	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	118390	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	120305	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	57921	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	54836	50.398	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.800%	
35) Dibromofluoromethane	5.391	113	40515	50.783	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.560%	
50) Toluene-d8	8.647	98	149413	51.848	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 103.700%	
62) 4-Bromofluorobenzene	11.079	95	58492	50.653	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.300%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	26064	57.491	ug/l	92
20) Methylene Chloride	2.788	84	9336	9.835	ug/l	91
43) Isopropyl Acetate	6.342	43	56469	26.089	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022324\
Data File : VX040439.D
Acq On : 23 Feb 2024 18:04
Operator : JC/MD
Sample : PB159217ZHE#09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB159217ZHE#09

Quant Time: Feb 23 22:23:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022224W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 23 06:47:29 2024
Response via : Initial Calibration

