

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022324\
 Data File : VX040441.D
 Acq On : 23 Feb 2024 18:51
 Operator : JC/MD
 Sample : PB159217ZHE#11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB159217ZHE#11

Manual Integrations
 APPROVED

Reviewed By : John Carlone 02/26/2024
 Supervised By : Mahesh Dadoda 02/26/2024

Quant Time: Feb 23 22:24:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022224W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 23 06:47:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	68744	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	122730	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127661	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	61097	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	57385	51.257	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 102.520%			
35) Dibromofluoromethane	5.385	113	41780	50.517	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.040%			
50) Toluene-d8	8.647	98	157815	52.827	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 105.660%			
62) 4-Bromofluorobenzene	11.079	95	62332	52.069	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 104.140%			
Target Compounds					Qvalue	
16) Acetone	2.380	43	27451	58.847	ug/l	100
20) Methylene Chloride	2.788	84	9445	9.670	ug/l	97
37) Ethyl Acetate	4.739	43	1881m	1.286	ug/l	
43) Isopropyl Acetate	6.342	43	57100	25.447	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022324\
Data File : VX040441.D
Acq On : 23 Feb 2024 18:51
Operator : JC/MD
Sample : PB159217ZHE#11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB159217ZHE#11

Manual Integrations
APPROVED

Reviewed By : John Carlone 02/26/2024
Supervised By : Mahesh Dadoda 02/26/2024

Quant Time: Feb 23 22:24:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022224W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 23 06:47:29 2024
Response via : Initial Calibration

