

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022420\
 Data File : VX014999.D
 Acq On : 24 Feb 2020 10:27
 Operator : JC/SP
 Sample : VX0224WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0224WBL01

Quant Time: Feb 25 00:51:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 10 12:09:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	453552	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	712147	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	656445	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	306260	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	257495	50.30	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.60%	
35) Dibromofluoromethane	5.48	113	217092	49.59	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.18%	
50) Toluene-d8	8.71	98	849569	50.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.68%	
62) 4-Bromofluorobenzene	11.13	95	294323	48.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.78%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022420\
Data File : VX014999.D
Acq On : 24 Feb 2020 10:27
Operator : JC/SP
Sample : VX0224WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0224WBL01

Quant Time: Feb 25 00:51:04 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 10 12:09:20 2020
Response via : Initial Calibration

