

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022422\
 Data File : VX027165.D
 Acq On : 24 Feb 2022 13:12
 Operator : JC/MD
 Sample : N1647-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 31382

Quant Time: Feb 25 00:03:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	95020	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	156215	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	139352	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	65594	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	56681	49.668	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	99.340%	
35) Dibromofluoromethane	5.391	113	49354	50.899	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	101.800%	
50) Toluene-d8	8.653	98	179419	49.486	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	98.980%	
62) 4-Bromofluorobenzene	11.079	95	67227	48.305	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	96.620%	
Target Compounds						
						Qvalue
16) Acetone	2.386	43	48656	97.108	ug/l	95
25) 2-Butanone	4.568	43	10895	13.654	ug/l	100
40) Benzene	6.050	78	12735	3.077	ug/l	99
51) 4-Methyl-2-Pentanone	8.574	43	13262	8.537	ug/l	98
52) Toluene	8.720	92	136516	51.517	ug/l	99
59) 2-Hexanone	9.439	43	1384	1.166	ug/l	96
67) Ethyl Benzene	10.195	91	47216	9.468	ug/l	99
68) m/p-Xylenes	10.299	106	85146	44.414	ug/l	97
69) o-Xylene	10.640	106	45031	23.404	ug/l	97
73) Isopropylbenzene	10.963	105	9843	2.097	ug/l	98
78) n-propylbenzene	11.305	91	35840	6.718	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	68325	17.312	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	305771	77.888	ug/l	97
85) sec-Butylbenzene	11.890	105	15006	3.118	ug/l	84
86) p-Isopropyltoluene	12.012	119	11686	2.849	ug/l	99
95) Naphthalene	13.774	128	73640	15.977	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022422\
Data File : VX027165.D
Acq On : 24 Feb 2022 13:12
Operator : JC/MD
Sample : N1647-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
31382

Quant Time: Feb 25 00:03:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

