

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022422\
 Data File : VX027168.D
 Acq On : 24 Feb 2022 14:23
 Operator : JC/MD
 Sample : N1659-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-3-MH

Quant Time: Feb 25 00:04:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	91574	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	150297	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	137563	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	66288	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	53568	48.707	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.420%
35) Dibromofluoromethane	5.391	113	46089	49.403	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.800%
50) Toluene-d8	8.653	98	174393	49.994	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.980%
62) 4-Bromofluorobenzene	11.079	95	67651	50.524	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.040%
Target Compounds						
					Qvalue	
16) Acetone	2.385	43	8608	17.826	ug/l	91
25) 2-Butanone	4.574	43	1115	1.450	ug/l	97
51) 4-Methyl-2-Pentanone	8.579	43	1875	1.254	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022422\
Data File : VX027168.D
Acq On : 24 Feb 2022 14:23
Operator : JC/MD
Sample : N1659-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-3-MH

Quant Time: Feb 25 00:04:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

