

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022422\
 Data File : VX027175.D
 Acq On : 24 Feb 2022 17:13
 Operator : JC/MD
 Sample : N1539-02 10X
 Misc : 5.31g/5.0mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PE-2

Quant Time: Feb 25 00:05:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 02/25/2022
 Supervised By : Mahesh Dadoda 02/25/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	93713	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	154593	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	142737	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	69282	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	54247	48.198	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	96.400%		
35) Dibromofluoromethane	5.391	113	46183	48.128	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.260%		
50) Toluene-d8	8.653	98	179393	49.998	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	100.000%		
62) 4-Bromofluorobenzene	11.085	95	69874	50.734	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	101.460%		
Target Compounds						
					Qvalue	
31) Cyclohexane	5.501	56	31810	18.872	ug/l #	52
39) Methylcyclohexane	7.385	83	54652	31.372	ug/l	99
52) Toluene	8.720	92	556250	212.115	ug/l	98
67) Ethyl Benzene	10.195	91	523657	102.521	ug/l	98
68) m/p-Xylenes	10.305	106	608743	310.006	ug/l	96
69) o-Xylene	10.646	106	349274	177.224	ug/l	95
73) Isopropylbenzene	10.963	105	54353	10.964	ug/l	99
78) n-propylbenzene	11.305	91	221813	39.366	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	350478	84.075	ug/l	98
84) 1,2,4-Trimethylbenzene	11.756	105	1095032	264.084	ug/l	96
86) p-Isopropyltoluene	12.012	119	11586m	2.674	ug/l	
95) Naphthalene	13.774	128	190409	39.112	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022422\
Data File : VX027175.D
Acq On : 24 Feb 2022 17:13
Operator : JC/MD
Sample : N1539-02 10X
Misc : 5.31g/5.0mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PE-2

Quant Time: Feb 25 00:05:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 02/25/2022
Supervised By : Mahesh Dadoda 02/25/2022

