

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX022422\
 Data File : VX027179.D
 Acq On : 24 Feb 2022 18:47
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 25 00:06:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	97	0.00
2 T	Dichlorodifluoromethane	50.000	49.356	1.3	97	0.00
3 P	Chloromethane	50.000	51.673	-3.3	99	0.00
4 C	Vinyl Chloride	50.000	48.928	2.1#	93	0.00
5 T	Bromomethane	50.000	53.071	-6.1	112	0.00
6 T	Chloroethane	50.000	48.694	2.6	96	0.00
7 T	Trichlorofluoromethane	50.000	49.209	1.6	95	0.00
8 T	Diethyl Ether	50.000	49.369	1.3	93	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	49.821	0.4	95	0.00
10 T	Methyl Iodide	50.000	54.376	-8.8	100	0.00
11 T	Tert butyl alcohol	250.000	225.989	9.6	88	-0.01
12 CM	1,1-Dichloroethene	50.000	50.760	-1.5#	97	0.00
13 T	Acrolein	250.000	241.769	3.3	97	0.00
14 T	Allyl chloride	50.000	49.784	0.4	94	0.00
15 T	Acrylonitrile	250.000	260.043	-4.0	97	0.00
16 T	Acetone	250.000	240.847	3.7	96	0.00
17 T	Carbon Disulfide	50.000	46.478	7.0	89	0.00
18 T	Methyl Acetate	50.000	49.173	1.7	93	0.00
19 T	Methyl tert-butyl Ether	50.000	51.603	-3.2	97	0.00
20 T	Methylene Chloride	50.000	51.121	-2.2	99	0.00
21 T	trans-1,2-Dichloroethene	50.000	51.630	-3.3	98	0.00
22 T	Diisopropyl ether	50.000	52.074	-4.1	98	0.00
23 T	Vinyl Acetate	250.000	263.812	-5.5	96	0.00
24 P	1,1-Dichloroethane	50.000	50.813	-1.6	97	0.00
25 T	2-Butanone	250.000	253.487	-1.4	96	0.00
26 T	2,2-Dichloropropane	50.000	48.346	3.3	90	0.00
27 T	cis-1,2-Dichloroethene	50.000	51.863	-3.7	98	0.00
28 T	Bromochloromethane	50.000	48.377	3.2	90	0.00
29 T	Tetrahydrofuran	250.000	252.490	-1.0	96	0.00
30 C	Chloroform	50.000	49.967	0.1#	95	0.00
31 T	Cyclohexane	50.000	49.462	1.1	95	0.00
32 T	1,1,1-Trichloroethane	50.000	50.508	-1.0	94	0.00
33 S	1,2-Dichloroethane-d4	50.000	47.762	4.5	95	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	98	0.00
35 S	Dibromofluoromethane	50.000	47.723	4.6	96	0.00
36 T	1,1-Dichloropropene	50.000	49.802	0.4	95	0.00
37 T	Ethyl Acetate	50.000	50.971	-1.9	99	0.00
38 T	Carbon Tetrachloride	50.000	49.371	1.3	93	0.00
39 T	Methylcyclohexane	50.000	51.436	-2.9	98	0.00
40 TM	Benzene	50.000	50.711	-1.4	97	0.00
41 T	Methacrylonitrile	50.000	51.869	-3.7	98	0.00
42 TM	1,2-Dichloroethane	50.000	50.454	-0.9	97	0.00
43 T	Isopropyl Acetate	50.000	51.997	-4.0	97	0.00
44 TM	Trichloroethene	50.000	50.727	-1.5	97	0.00
45 C	1,2-Dichloropropane	50.000	53.440	-6.9#	99	0.00
46 T	Dibromomethane	50.000	51.386	-2.8	96	0.00
47 T	Bromodichloromethane	50.000	50.265	-0.5	96	0.00
48 T	Methyl methacrylate	50.000	53.770	-7.5	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022422\
 Data File : VX027179.D
 Acq On : 24 Feb 2022 18:47
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 25 00:06:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	920.158	8.0	87	0.00
50 S	Toluene-d8	50.000	47.052	5.9	94	0.00
51 T	4-Methyl-2-Pentanone	250.000	264.148	-5.7	99	0.00
52 CM	Toluene	50.000	51.013	-2.0#	97	0.00
53 T	t-1,3-Dichloropropene	50.000	46.338	7.3	95	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.713	-5.4	96	0.00
55 T	1,1,2-Trichloroethane	50.000	52.571	-5.1	99	0.00
56 T	Ethyl methacrylate	50.000	53.048	-6.1	97	0.00
57 T	1,3-Dichloropropane	50.000	53.020	-6.0	99	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	266.675	-6.7	99	0.00
59 T	2-Hexanone	250.000	259.845	-3.9	97	0.00
60 T	Dibromochloromethane	50.000	51.108	-2.2	94	0.00
61 T	1,2-Dibromoethane	50.000	51.779	-3.6	96	0.00
62 S	4-Bromofluorobenzene	50.000	47.718	4.6	94	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	99	0.00
64 T	Tetrachloroethene	50.000	50.366	-0.7	97	0.00
65 PM	Chlorobenzene	50.000	50.638	-1.3	98	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.759	-1.5	97	0.00
67 C	Ethyl Benzene	50.000	51.006	-2.0#	97	0.00
68 T	m/p-Xylenes	100.000	104.290	-4.3	98	0.00
69 T	o-Xylene	50.000	51.326	-2.7	98	0.00
70 T	Styrene	50.000	52.875	-5.8	98	0.00
71 P	Bromoform	50.000	43.708	12.6	91	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	95	0.00
73 T	Isopropylbenzene	50.000	52.179	-4.4	97	0.00
74 T	N-ethyl acetate	50.000	54.684	-9.4	96	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	51.122	-2.2	97	0.00
76 T	1,2,3-Trichloropropane	50.000	53.570	-7.1	99	0.00
77 T	Bromobenzene	50.000	51.493	-3.0	96	0.00
78 T	n-propylbenzene	50.000	52.270	-4.5	96	0.00
79 T	2-Chlorotoluene	50.000	51.669	-3.3	97	0.00
80 T	1,3,5-Trimethylbenzene	50.000	52.660	-5.3	96	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	43.986	12.0	88	0.00
82 T	4-Chlorotoluene	50.000	52.155	-4.3	97	0.00
83 T	tert-Butylbenzene	50.000	52.393	-4.8	97	0.00
84 T	1,2,4-Trimethylbenzene	50.000	53.406	-6.8	98	0.00
85 T	sec-Butylbenzene	50.000	53.731	-7.5	98	0.00
86 T	p-Isopropyltoluene	50.000	52.970	-5.9	96	0.00
87 T	1,3-Dichlorobenzene	50.000	51.400	-2.8	96	0.00
88 T	1,4-Dichlorobenzene	50.000	49.952	0.1	95	0.00
89 T	n-Butylbenzene	50.000	51.745	-3.5	93	0.00
90 T	Hexachloroethane	50.000	49.052	1.9	88	0.00
91 T	1,2-Dichlorobenzene	50.000	50.736	-1.5	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	49.868	0.3	90	0.00
93 T	1,2,4-Trichlorobenzene	50.000	51.619	-3.2	94	0.00
94 T	Hexachlorobutadiene	50.000	50.099	-0.2	92	0.00
95 T	Naphthalene	50.000	53.178	-6.4	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022422\
Data File : VX027179.D
Acq On : 24 Feb 2022 18:47
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Feb 25 00:06:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	50.869	-1.7	93	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6