



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022422\
Data File : VX027159.D
Acq On : 24 Feb 2022 09:44
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Title : SW846 8260
Last Update : Wed Feb 23 13:06:57 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	17.0	4427	PASS
75	95	30	60	49.2	12826	PASS
95	95	100	100	100.0	26067	PASS
96	95	5	9	6.7	1743	PASS
173	174	0.00	2	1.0	199	PASS
174	95	50	100	74.6	19448	PASS
175	174	5	9	7.6	1473	PASS
176	174	95	101	96.8	18821	PASS
177	176	5	9	6.7	1257	PASS