

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022423\
 Data File : VX034260.D
 Acq On : 24 Feb 2023 12:52
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Feb 24 22:21:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	280572	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	456454	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	400417	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	174683	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	160910	42.584	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.160%
35) Dibromofluoromethane	5.385	113	141611	47.649	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.300%
50) Toluene-d8	8.646	98	542593	50.851	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.700%
62) 4-Bromofluorobenzene	11.079	95	182200	44.641	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.280%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.788	84	2152	0.613	ug/l #	90
64) Tetrachloroethene	9.280	164	584	0.207	ug/l #	74
86) p-Isopropyltoluene	12.012	119	2318	0.201	ug/l #	76
87) 1,3-Dichlorobenzene	11.975	146	1513	0.252	ug/l	91
89) n-Butylbenzene	12.335	91	2465	0.240	ug/l	94
93) 1,2,4-Trichlorobenzene	13.591	180	1783	0.472	ug/l	95
94) Hexachlorobutadiene	13.725	225	1039	0.664	ug/l	98
95) Naphthalene	13.780	128	4368	0.344	ug/l	98
96) 1,2,3-Trichlorobenzene	13.963	180	1582	0.422	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022423\
Data File : VX034260.D
Acq On : 24 Feb 2023 12:52
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Feb 24 22:21:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

