

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022423\
 Data File : VX034263.D
 Acq On : 24 Feb 2023 14:02
 Operator : JC/MD
 Sample : 01674-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Feb 24 22:22:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	272592	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	451454	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	394134	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	174305	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	160750	43.787	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.580%
35) Dibromofluoromethane	5.385	113	138470	47.108	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.220%
50) Toluene-d8	8.647	98	530854	50.301	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.600%
62) 4-Bromofluorobenzene	11.079	95	181068	44.855	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.700%
Target Compounds						
20) Methylene Chloride	2.800	84	887	0.260	ug/l	Qvalue # 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022423\
Data File : VX034263.D
Acq On : 24 Feb 2023 14:02
Operator : JC/MD
Sample : 01674-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Feb 24 22:22:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

