



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022423\
Data File : VX034254.D
Acq On : 24 Feb 2023 09:14
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Title : SW846 8260
Last Update : Fri Feb 17 12:33:35 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.3	16727	PASS
75	95	30	60	49.6	42971	PASS
95	95	100	100	100.0	86683	PASS
96	95	5	9	7.0	6085	PASS
173	174	0.00	2	0.9	660	PASS
174	95	50	100	81.5	70648	PASS
175	174	5	9	7.4	5216	PASS
176	174	95	101	97.1	68616	PASS
177	176	5	9	6.4	4409	PASS