

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022425\
 Data File : VX045025.D
 Acq On : 24 Feb 2025 12:20
 Operator : JC/MD
 Sample : VX0224WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0224WBL01

Quant Time: Feb 25 00:30:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X022125W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Feb 22 01:06:36 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.891	128	19274	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.757	114	100345	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.049	117	86987	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.946	65	55082	29.424	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	98.067%
60) 4-Bromofluorobenzene	11.079	95	45953	28.263	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	94.200%
63) Toluene-d8	8.646	98	149824	31.134	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	103.767%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022425\
Data File : VX045025.D
Acq On : 24 Feb 2025 12:20
Operator : JC/MD
Sample : VX0224WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0224WBL01

Quant Time: Feb 25 00:30:23 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X022125W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Feb 22 01:06:36 2025
Response via : Initial Calibration

