

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022519\
 Data File : VX007705.D
 Acq On : 26 Feb 2019 09:34
 Operator : JC/SP
 Sample : K1348-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OK-03-022219-A


Quant Time: Feb 26 11:57:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 01:53:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	326564	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	560952	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	557062	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	246942	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	204371	50.52	ug/l	0.00
Spiked Amount			Recovery	=	101.04%	
35) Dibromofluoromethane	5.51	113	189333	50.49	ug/l	0.00
Spiked Amount			Recovery	=	100.98%	
50) Toluene-d8	8.71	98	708679	49.37	ug/l	0.00
Spiked Amount			Recovery	=	98.74%	
62) 4-Bromofluorobenzene	11.13	95	245735	46.64	ug/l	0.00
Spiked Amount			Recovery	=	93.28%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
5) Bromomethane	1.65	94	232	Below Cal	#	3
6) Chloroethane	1.59	64	5569	2.549	ug/l #	50
8) Diethyl Ether	2.19	74	1040	0.483	ug/l	81
11) Tert butyl alcohol	3.06	59	1597	2.081	ug/l	98
13) Acrolein	2.31	56	500	0.733	ug/l #	53
16) Acetone	2.45	43	23558	13.096	ug/l	98
18) Methyl Acetate	2.78	43	1552	0.395	ug/l #	86
20) Methylene Chloride	2.86	84	436697	131.471	ug/l	94
25) 2-Butanone	4.71	43	6361	2.321	ug/l	99
29) Tetrahydrofuran	5.16	42	2089	1.157	ug/l #	75
36) 1,1-Dichloropropene	5.66	75	22853	4.531	ug/l #	51
38) Carbon Tetrachloride	5.66	117	30483	6.017	ug/l #	16
43) Isopropyl Acetate	6.46	43	17553	2.036	ug/l #	90
47) Bromodichloromethane	7.90	83	1586	0.325	ug/l #	50
48) Methyl methacrylate	7.68	41	8954	1.981	ug/l #	12
49) 1,4-Dioxane	7.75	88	316	3.237	ug/l #	25
51) 4-Methyl-2-Pentanone	8.62	43	89704	15.543	ug/l #	52
53) t-1,3-Dichloropropene	9.02	75	230	2.838	ug/l #	43
54) cis-1,3-Dichloropropene	8.62	75	43777	7.399	ug/l #	66
59) 2-Hexanone	9.54	43	130356	28.769	ug/l #	53
67) Ethyl Benzene	10.25	91	6838	0.343	ug/l	99
68) m/p-Xylenes	10.36	106	6171	0.795	ug/l	88
76) 1,2,3-Trichloropropane	11.13	75	117815	22.383	ug/l #	33
78) n-propylbenzene	11.36	91	4807	0.241	ug/l	91
80) 1,3,5-Trimethylbenzene	11.50	105	6443	0.444	ug/l	96
81) trans-1,4-Dichloro-2-buten	11.13	75	117815	59.378	ug/l #	9
84) 1,2,4-Trimethylbenzene	11.80	105	17159	1.155	ug/l	100
95) Naphthalene	13.83	128	9007	0.554	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. -0.00 min

Lab File: VX007705.D

Acq: 26 Feb 2019 09:34

Instrument :

MSVOA_X

ClientSampleId :


OK-03-022219-A

Tgt Ion:168 Resp: 326564

Ion Ratio Lower Upper




168 100

99 54.4 37.9 56.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#5

Bromomethane

Concen: Below Cal

RT: 1.65 min Scan# 81

Delta R.T. 0.01 min

Lab File: VX007705.D


Acq: 26 Feb 2019 09:34

Tgt Ion: 94 Resp: 232

Ion Ratio Lower Upper



94 100


96 0.0 73.9 110.9#



Abundance Ion 93.90 (93.60 to 94.60): VX0

Ion 95.90 (95.60 to 96.60): VX0





#6

Chloroethane

Concen: 2.549 ug/l

RT: 1.59 min Scan# 72

Delta R.T. -0.13 min

Lab File: VX007705.D

Acq: 26 Feb 2019 09:34

Instrument :

MSVOA_X

ClientSampleId :


OK-03-022219-A

Tgt Ion: 64 Resp: 5569

Ion Ratio Lower Upper

64 100

66 4.3 25.5 38.3#



Abundance Ion 63.90 (63.60 to 64.60): VX0

Ion 65.90 (65.60 to 66.60): VX0

1.59

8000



6000

4000

2000

0

Time--> 1.55 1.60



#8

Diethyl Ether

Concen: 0.483 ug/l

RT: 2.19 min Scan# 170

Delta R.T. -0.00 min

Lab File: VX007705.D


Acq: 26 Feb 2019 09:34

Tgt Ion: 74 Resp: 1040

Ion Ratio Lower Upper

74 100

45 109.3 45.7 137.1



Abundance Ion 74.00 (73.70 to 74.70): VX0

Ion 45.00 (44.70 to 45.70): VX0

1000

800


600


400

200

0

Time--> 2.15 2.20









#11
Tert butyl alcohol
Concen: 2.081 ug/l
RT: 3.06 min Scan# 312
Delta R.T. -0.02 min
Lab File: VX007705.D
Acq: 26 Feb 2019 09:34




Instrument :
MSVOA_X
ClientSampleId :
OK-03-022219-A


Tgt Ion: 59 Resp: 1597
Ion Ratio Lower Upper
59 100
57 10.9 8.1 12.1



#13
Acrolein
Concen: 0.733 ug/l
RT: 2.31 min Scan# 189
Delta R.T. 0.01 min
Lab File: VX007705.D
Acq: 26 Feb 2019 09:34

Tgt Ion: 56 Resp: 500
Ion Ratio Lower Upper
56 100
55 109.6 56.9 85.3#





#16

Acetone

Concen: 13.096 ug/l

RT: 2.45 min Scan# 212

Delta R.T. -0.00 min

Lab File: VX007705.D

Acq: 26 Feb 2019 09:34

Instrument :

MSVOA_X

ClientSampleId :





OK-03-022219-A

Tgt Ion: 43 Resp: 23558

Ion Ratio Lower Upper

43 100

58 30.1 24.9 37.3



#18

Methyl Acetate

Concen: 0.395 ug/l

RT: 2.78 min Scan# 267

Delta R.T. -0.00 min

Lab File: VX007705.D




Acq: 26 Feb 2019 09:34


Tgt Ion: 43 Resp: 1552

Ion Ratio Lower Upper

43 100

74 31.1 19.4 29.2#









#20
Methylene Chloride
Concen: 131.471 ug/l
RT: 2.86 min Scan# 280
Delta R.T. -0.00 min
Lab File: VX007705.D
Acq: 26 Feb 2019 09:34




Instrument :
MSVOA_X
ClientSampleId :
OK-03-022219-A


Tgt Ion:	84	Resp:	436697
Ion	Ratio	Lower	Upper
84	100		
49	128.0	94.8	142.2
51	39.0	29.8	44.8
86	64.6	54.1	81.1



#25
2-Butanone
Concen: 2.321 ug/l
RT: 4.71 min Scan# 583
Delta R.T. 0.01 min
Lab File: VX007705.D
Acq: 26 Feb 2019 09:34

Tgt Ion:	43	Resp:	6361
Ion	Ratio	Lower	Upper
43	100		
72	27.3	22.2	33.2





#29

Tetrahydrofuran

Concen: 1.157 ug/l

RT: 5.16 min Scan# 657

Delta R.T. 0.01 min

Lab File: VX007705.D

Acq: 26 Feb 2019 09:34

Instrument :

MSVOA_X

ClientSampleId :

OK-03-022219-A





Tgt Ion: 42 Resp: 2089

Ion Ratio Lower Upper

42 100

72 22.8 37.6 56.4#

71 34.8 34.6 51.8



#33

1,2-Dichloroethane-d4

Concen: 50.523 ug/l

RT: 6.07 min Scan# 806

Delta R.T. -0.00 min

Lab File: VX007705.D




Acq: 26 Feb 2019 09:34


Tgt Ion: 65 Resp: 204371

Ion Ratio Lower Upper

65 100

67 52.6 0.0 108.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.86 min Scan# 936

Delta R.T. -0.00 min

Lab File: VX007705.D

Acq: 26 Feb 2019 09:34

Instrument :

MSVOA_X

ClientSampleId :

OK-03-022219-A


Tgt Ion:114 Resp: 560952

Ion Ratio Lower Upper

114 100

63 19.3 0.0 35.6




88 15.0 0.0 29.6



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 50.492 ug/l

RT: 5.51 min Scan# 714

Delta R.T. 0.01 min

Lab File: VX007705.D

Acq: 26 Feb 2019 09:34


Tgt Ion:113 Resp: 189333

Ion Ratio Lower Upper

113 100

111 102.2 81.8 122.8



192 20.1 18.5 27.7




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#36
 1,1-Dichloropropene
 Concen: 4.531 ug/l
 RT: 5.66 min Scan# 739
 Delta R.T. -0.14 min
 Lab File: VX007705.D
 Acq: 26 Feb 2019 09:34




Instrument :
 MSVOA_X
 ClientSampleId :
 OK-03-022219-A


Tgt Ion: 75 Resp: 22853
 Ion Ratio Lower Upper
 75 100
 110 11.0 20.0 59.9#
 77 2.0 24.7 37.1#



#38
 Carbon Tetrachloride
 Concen: 6.017 ug/l
 RT: 5.66 min Scan# 739
 Delta R.T. -0.13 min
 Lab File: VX007705.D
 Acq: 26 Feb 2019 09:34

Tgt Ion: 117 Resp: 30483
 Ion Ratio Lower Upper
 117 100
 119 3.9 75.3 112.9#
 121 0.0 25.0 37.4#





#43

Isopropyl Acetate

Concen: 2.036 ug/l

RT: 6.46 min Scan# 871

Delta R.T. 0.01 min

Lab File: VX007705.D

Acq: 26 Feb 2019 09:34





Instrument :

MSVOA_X

ClientSampleId :

OK-03-022219-A

Tgt Ion:	43	Resp:	17553
Ion	Ratio	Lower	Upper
43	100		
61	25.3	16.1	24.1#
87	18.0	11.5	17.3#



#47

Bromodichloromethane

Concen: 0.325 ug/l




RT: 7.90 min Scan# 1107


Delta R.T. 0.01 min

Lab File: VX007705.D

Acq: 26 Feb 2019 09:34

Tgt Ion:	83	Resp:	1586
Ion	Ratio	Lower	Upper
83	100		
85	22.6	51.4	77.0#
127	0.0	8.1	12.1#





#48

Methyl methacrylate

Concen: 1.981 ug/l

RT: 7.68 min Scan# 1070

Delta R.T. -0.10 min

Lab File: VX007705.D

Acq: 26 Feb 2019 09:34

Instrument :

MSVOA_X

ClientSampleId :

OK-03-022219-A





Tgt Ion: 41 Resp: 8954

Ion Ratio Lower Upper

41 100

69 1.0 70.2 105.4#

39 0.0 47.4 71.0#



#49

1,4-Dioxane

Concen: 3.237 ug/l

RT: 7.75 min Scan# 1082

Delta R.T. -0.00 min

Lab File: VX007705.D

Acq: 26 Feb 2019 09:34




Tgt Ion: 88 Resp: 316


Ion Ratio Lower Upper

88 100

43 0.0 24.2 36.4#

58 0.0 53.0 79.4#





#50

Toluene-d8

Concen: 49.366 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. -0.00 min

Lab File: VX007705.D

Acq: 26 Feb 2019 09:34

Instrument :

MSVOA_X

ClientSampleId :


OK-03-022219-A

Tgt Ion: 98 Resp: 708679

Ion Ratio Lower Upper




98 100

100 64.1 51.6 77.4



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX



#51

4-Methyl-2-Pentanone

Concen: 15.543 ug/l

RT: 8.62 min Scan# 1224

Delta R.T. -0.03 min

Lab File: VX007705.D


Acq: 26 Feb 2019 09:34

Tgt Ion: 43 Resp: 89704

Ion Ratio Lower Upper



43 100


58 9.9 31.2 46.8#



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0





#53

t-1,3-Dichloropropene

Concen: 2.838 ug/l

RT: 9.02 min Scan# 1291

Delta R.T. -0.02 min

Lab File: VX007705.D

Acq: 26 Feb 2019 09:34

Instrument :

MSVOA_X

ClientSampleId :





OK-03-022219-A

Tgt Ion: 75 Resp: 230

Ion Ratio Lower Upper

75 100

77 0.0 25.2 37.8#



#54

cis-1,3-Dichloropropene

Concen: 7.399 ug/l

RT: 8.62 min Scan# 1224

Delta R.T. 0.18 min

Lab File: VX007705.D

Acq: 26 Feb 2019 09:34




Tgt Ion: 75 Resp: 43777


Ion Ratio Lower Upper

75 100

77 0.9 26.2 39.4#

39 34.4 37.1 55.7#





#59

2-Hexanone

Concen: 28.769 ug/l

RT: 9.54 min Scan# 1376

Delta R.T. 0.05 min

Lab File: VX007705.D

Acq: 26 Feb 2019 09:34

Instrument :

MSVOA_X

ClientSampleId :





OK-03-022219-A

Tgt Ion: 43 Resp: 130356

Ion Ratio Lower Upper

43 100

58 20.1 27.1 81.3#



#62

4-Bromofluorobenzene

Concen: 46.638 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. -0.00 min

Lab File: VX007705.D

Acq: 26 Feb 2019 09:34




Tgt Ion: 95 Resp: 245735


Ion Ratio Lower Upper

95 100

174 84.3 0.0 189.2

176 81.4 0.0 186.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. -0.00 min
Lab File: VX007705.D
Acq: 26 Feb 2019 09:34

Instrument :
MSVOA_X
ClientSampleId :
OK-03-022219-A

Tgt Ion:117 Resp: 557062
Ion Ratio Lower Upper
117 100
82 52.8 42.1 63.1
119 32.7 26.5 39.7




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V






#67
Ethyl Benzene
Concen: 0.343 ug/l
RT: 10.25 min Scan# 1492
Delta R.T. -0.00 min
Lab File: VX007705.D
Acq: 26 Feb 2019 09:34

Tgt Ion: 91 Resp: 6838
Ion Ratio Lower Upper
91 100
106 32.0 25.4 38.0



Abundance Ion 91.00 (90.70 to 91.70): VX0
Ion 106.00 (105.70 to 106.70): V





#68

m/p-Xylenes

Concen: 0.795 ug/l

RT: 10.36 min Scan# 1510

Delta R.T. -0.00 min

Lab File: VX007705.D

Acq: 26 Feb 2019 09:34

Instrument :

MSVOA_X

ClientSampleId :


OK-03-022219-A

Tgt Ion:106 Resp: 6171

Ion Ratio Lower Upper




106 100

91 177.7 156.6 234.8



Abundance Ion 106.00 (105.70 to 106.70): V

Ion 91.00 (90.70 to 91.70): VX0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.08 min Scan# 1792

Delta R.T. -0.00 min

Lab File: VX007705.D

Acq: 26 Feb 2019 09:34


Tgt Ion:152 Resp: 246942

Ion Ratio Lower Upper

152 100

115 56.4 38.0 114.0



150 154.9 0.0 346.4




Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V









#76
 1,2,3-Trichloropropane
 Concen: 22.383 ug/l
 RT: 11.13 min Scan# 1637
 Delta R.T. -0.16 min
 Lab File: VX007705.D
 Acq: 26 Feb 2019 09:34




Instrument :
 MSVOA_X
 ClientSampleId :
 OK-03-022219-A


Tgt Ion: 75 Resp: 117815
 Ion Ratio Lower Upper
 75 100
 77 1.4 23.0 69.0#



#78
 n-propylbenzene
 Concen: 0.241 ug/l
 RT: 11.36 min Scan# 1674
 Delta R.T. -0.00 min
 Lab File: VX007705.D
 Acq: 26 Feb 2019 09:34

Tgt Ion: 91 Resp: 4807
 Ion Ratio Lower Upper
 91 100
 120 29.3 12.4 37.2





#80

1,3,5-Trimethylbenzene

Concen: 0.444 ug/l

RT: 11.50 min Scan# 1697

Delta R.T. -0.01 min

Lab File: VX007705.D

Acq: 26 Feb 2019 09:34

Instrument :

MSVOA_X

ClientSampleId :


OK-03-022219-A

Tgt Ion:105 Resp: 6443

Ion Ratio Lower Upper




105 100

120 53.4 25.3 75.8



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V



#81

trans-1,4-Dichloro-2-butene

Concen: 59.378 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. 0.06 min

Lab File: VX007705.D

Acq: 26 Feb 2019 09:34


Tgt Ion: 75 Resp: 117815

Ion Ratio Lower Upper

75 100

53 0.5 81.0 121.6#



89 0.0 41.4 62.0#




Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 53.00 (52.70 to 53.70): VX0

Ion 88.90 (88.60 to 89.60): VX0









#84
 1,2,4-Trimethylbenzene
 Concen: 1.155 ug/l
 RT: 11.80 min Scan# 1747
 Delta R.T. -0.00 min
 Lab File: VX007705.D
 Acq: 26 Feb 2019 09:34

Instrument :
 MSVOA_X
 ClientSampleId :
 OK-03-022219-A

Tgt Ion:105 Resp: 17159
 Ion Ratio Lower Upper
 105 100
 120 46.1 23.1 69.3



#95
 Naphthalene
 Concen: 0.554 ug/l
 RT: 13.83 min Scan# 2080
 Delta R.T. -0.00 min
 Lab File: VX007705.D
 Acq: 26 Feb 2019 09:34

Tgt Ion:128 Resp: 9007
 Ion Ratio Lower Upper
 128 100
 127 16.5 10.5 15.7#
 129 13.2 8.7 13.1#

