

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022519\
 Data File : VX007672.D
 Acq On : 25 Feb 2019 14:34
 Operator : JC/SP
 Sample : K1578-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB201-EB-20190221

Quant Time: Feb 26 04:11:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 01:53:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	374129	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	637056	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	585487	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	263450	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	226865	48.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.90%	
35) Dibromofluoromethane	5.50	113	204939	48.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.24%	
50) Toluene-d8	8.71	98	758821	46.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.08%	
62) 4-Bromofluorobenzene	11.13	95	262525	43.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.74%	
Target Compounds						
16) Acetone	2.45	43	13865	6.728	ug/l	96
52) Toluene	8.78	92	7065	0.634	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022519\
Data File : VX007672.D
Acq On : 25 Feb 2019 14:34
Operator : JC/SP
Sample : K1578-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB201-EB-20190221

Quant Time: Feb 26 04:11:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 01:53:13 2019
Response via : Initial Calibration

