

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022519\
 Data File : VX007678.D
 Acq On : 25 Feb 2019 17:13
 Operator : JC/SP
 Sample : K1578-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB201-GW-240-242

Quant Time: Feb 26 10:00:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 01:53:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	355441	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	609533	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	565840	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	253090	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	213185	48.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.84%	
35) Dibromofluoromethane	5.50	113	197501	48.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.94%	
50) Toluene-d8	8.71	98	726919	46.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.20%	
62) 4-Bromofluorobenzene	11.13	95	254081	44.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.76%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluoroet	2.39	101	36319	10.530	ug/l	92
12) 1,1-Dichloroethene	2.38	96	24373	7.839	ug/l	98
16) Acetone	2.45	43	8148	4.162	ug/l	97
24) 1,1-Dichloroethane	3.70	63	4613	0.709	ug/l #	90
27) cis-1,2-Dichloroethene	4.60	96	3702	0.861	ug/l	84
44) Trichloroethene	7.21	130	354939	76.274	ug/l	96
64) Tetrachloroethene	9.34	164	72434	15.769	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022519\
Data File : VX007678.D
Acq On : 25 Feb 2019 17:13
Operator : JC/SP
Sample : K1578-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB201-GW-240-242

Quant Time: Feb 26 10:00:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 01:53:13 2019
Response via : Initial Calibration

