

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022519\
 Data File : VX007680.D
 Acq On : 25 Feb 2019 18:07
 Operator : JC/SP
 Sample : K1578-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB201-GW-280-282

Quant Time: Feb 26 04:25:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 01:53:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	365316	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	611610	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	580057	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	269138	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	220066	48.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.26%	
35) Dibromofluoromethane	5.51	113	198835	48.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.26%	
50) Toluene-d8	8.71	98	741751	47.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.78%	
62) 4-Bromofluorobenzene	11.13	95	262023	45.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.22%	
Target Compounds						
7) Trichlorofluoromethane	1.94	101	10446	1.781	ug/l	97
9) 1,1,2-Trichlorotrifluoroet	2.39	101	17041	4.807	ug/l	89
12) 1,1-Dichloroethene	2.39	96	7012	2.194	ug/l	90
16) Acetone	2.45	43	11608	5.768	ug/l	100
24) 1,1-Dichloroethane	3.70	63	2797	0.418	ug/l #	56
27) cis-1,2-Dichloroethene	4.60	96	9346	2.114	ug/l	88
44) Trichloroethene	7.21	130	152608	32.683	ug/l	94
64) Tetrachloroethene	9.34	164	134201	28.499	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022519\
Data File : VX007680.D
Acq On : 25 Feb 2019 18:07
Operator : JC/SP
Sample : K1578-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB201-GW-280-282

Quant Time: Feb 26 04:25:28 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 01:53:13 2019
Response via : Initial Calibration

