

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022519\
 Data File : VX007681.D
 Acq On : 25 Feb 2019 18:33
 Operator : JC/SP
 Sample : K1578-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB201-GW-300-302

Quant Time: Feb 26 04:39:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 01:53:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	335307	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	575532	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	536921	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	242572	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	205556	49.49	ug/l	0.00
Spiked Amount			Recovery	=	98.98%	
35) Dibromofluoromethane	5.51	113	189791	49.33	ug/l	0.00
Spiked Amount			Recovery	=	98.66%	
50) Toluene-d8	8.71	98	688718	46.76	ug/l	0.00
Spiked Amount			Recovery	=	93.52%	
62) 4-Bromofluorobenzene	11.13	95	233768	43.24	ug/l	0.00
Spiked Amount			Recovery	=	86.48%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.20	85	4604	1.491	ug/l	99
7) Trichlorofluoromethane	1.94	101	85540	15.889	ug/l	100
9) 1,1,2-Trichlorotrifluoroet	2.39	101	4427	1.361	ug/l #	87
12) 1,1-Dichloroethene	2.38	96	27459	9.362	ug/l	98
16) Acetone	2.45	43	8672	4.695	ug/l	99
19) Methyl tert-butyl Ether	3.20	73	8809	0.862	ug/l	96
24) 1,1-Dichloroethane	3.70	63	7171	1.169	ug/l #	86
27) cis-1,2-Dichloroethene	4.60	96	2582	0.636	ug/l	74
30) Chloroform	5.22	83	14716	2.239	ug/l #	82
44) Trichloroethene	7.21	130	54841	12.481	ug/l	93
64) Tetrachloroethene	9.34	164	86526	19.851	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022519\
Data File : VX007681.D
Acq On : 25 Feb 2019 18:33
Operator : JC/SP
Sample : K1578-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB201-GW-300-302

Quant Time: Feb 26 04:39:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 01:53:13 2019
Response via : Initial Calibration

