

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022519\
 Data File : VX007688.D
 Acq On : 26 Feb 2019 02:02
 Operator : JC/SP
 Sample : VX0225WBL02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0225WBL02

Quant Time: Feb 26 04:49:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 01:53:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	359575	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	607440	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	576278	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	264543	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	215967	48.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.98%	
35) Dibromofluoromethane	5.51	113	196945	48.50	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.00%	
50) Toluene-d8	8.71	98	739569	47.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.16%	
62) 4-Bromofluorobenzene	11.13	95	257782	45.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.36%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022519\
Data File : VX007688.D
Acq On : 26 Feb 2019 02:02
Operator : JC/SP
Sample : VX0225WBL02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0225WBL02

Quant Time: Feb 26 04:49:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 01:53:13 2019
Response via : Initial Calibration

