

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022519\
 Data File : VX007694.D
 Acq On : 26 Feb 2019 04:41
 Operator : JC/SP
 Sample : K1603-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 M1-T1-OZ(PRE)

Quant Time: Feb 26 05:35:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 01:53:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	350796	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	591881	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	559855	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	255120	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	209815	48.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.58%	
35) Dibromofluoromethane	5.50	113	196923	49.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.54%	
50) Toluene-d8	8.71	98	717255	47.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.70%	
62) 4-Bromofluorobenzene	11.13	95	253853	45.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.32%	
Target Compounds						
19) Methyl tert-butyl Ether	3.21	73	15812	1.480	ug/l #	88
24) 1,1-Dichloroethane	3.70	63	8807	1.372	ug/l	97
29) Tetrahydrofuran	5.16	42	8597	4.431	ug/l	95
44) Trichloroethene	7.21	130	31365	6.941	ug/l	97
64) Tetrachloroethene	9.34	164	14866	3.271	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022519\
Data File : VX007694.D
Acq On : 26 Feb 2019 04:41
Operator : JC/SP
Sample : K1603-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
M1-T1-OZ(PRE)

Quant Time: Feb 26 05:35:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 01:53:13 2019
Response via : Initial Calibration

