

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022519\
 Data File : VX007695.D
 Acq On : 26 Feb 2019 05:08
 Operator : JC/SP
 Sample : K1603-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 M1-T2-OZ(PRE)

Quant Time: Feb 26 05:41:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 01:53:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	352505	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	596819	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	561451	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	255945	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	215123	49.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.54%	
35) Dibromofluoromethane	5.50	113	199019	49.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.78%	
50) Toluene-d8	8.71	98	715997	46.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.76%	
62) 4-Bromofluorobenzene	11.13	95	251392	44.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.68%	
Target Compounds						
					Qvalue	
19) Methyl tert-butyl Ether	3.21	73	5265	0.490	ug/l	91
24) 1,1-Dichloroethane	3.70	63	7343	1.138	ug/l	98
44) Trichloroethene	7.21	130	5990	1.315	ug/l	97
64) Tetrachloroethene	9.34	164	4501	0.988	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022519\
Data File : VX007695.D
Acq On : 26 Feb 2019 05:08
Operator : JC/SP
Sample : K1603-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
M1-T2-OZ(PRE)

Quant Time: Feb 26 05:41:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 01:53:13 2019
Response via : Initial Calibration

