

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX022519\
 Data File : VX007697.D
 Acq On : 26 Feb 2019 06:01
 Operator : JC/SP
 Sample : K1603-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 M1-T0-UV(PRE)

Quant Time: Feb 26 06:45:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 01:53:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	344665	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	582574	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	550228	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	250354	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	207483	48.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.20%	
35) Dibromofluoromethane	5.51	113	194930	50.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.10%	
50) Toluene-d8	8.71	98	710993	47.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.38%	
62) 4-Bromofluorobenzene	11.13	95	244848	44.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.48%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.38	96	2704	0.897	ug/l #	85
19) Methyl tert-butyl Ether	3.21	73	35345	3.367	ug/l	99
24) 1,1-Dichloroethane	3.70	63	8824	1.399	ug/l	94
27) cis-1,2-Dichloroethene	4.60	96	16262	3.899	ug/l	92
44) Trichloroethene	7.21	130	331316	74.492	ug/l	97
64) Tetrachloroethene	9.34	164	36011	8.062	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022519\
Data File : VX007697.D
Acq On : 26 Feb 2019 06:01
Operator : JC/SP
Sample : K1603-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
M1-T0-UV(PRE)

Quant Time: Feb 26 06:45:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 01:53:13 2019
Response via : Initial Calibration

