

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022519\
 Data File : VX007700.D
 Acq On : 26 Feb 2019 07:21
 Operator : JC/SP
 Sample : K1603-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 M1-T3-UV(PRE)

Quant Time: Feb 26 11:35:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 01:53:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	343432	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	586478	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	549162	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	248475	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	203613	47.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.72%	
35) Dibromofluoromethane	5.51	113	191574	48.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.74%	
50) Toluene-d8	8.71	98	707448	47.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.28%	
62) 4-Bromofluorobenzene	11.13	95	248118	45.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.08%	
Target Compounds						
16) Acetone	2.45	43	6334	3.348	ug/l	96
24) 1,1-Dichloroethane	3.70	63	5457	0.868	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022519\
Data File : VX007700.D
Acq On : 26 Feb 2019 07:21
Operator : JC/SP
Sample : K1603-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
M1-T3-UV(PRE)

Quant Time: Feb 26 11:35:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 01:53:13 2019
Response via : Initial Calibration

