

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022519\
 Data File : VX007702.D
 Acq On : 26 Feb 2019 08:14
 Operator : JC/SP
 Sample : K1603-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 M2-T3-UV(PRE)

Quant Time: Feb 26 11:50:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 01:53:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	339762	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	574889	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	541957	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	244050	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	206536	49.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.16%	
35) Dibromofluoromethane	5.50	113	191440	49.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.64%	
50) Toluene-d8	8.71	98	700659	47.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.24%	
62) 4-Bromofluorobenzene	11.13	95	240554	44.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.10%	
Target Compounds						
16) Acetone	2.45	43	6098	3.258	ug/l	97
29) Tetrahydrofuran	5.15	42	9654	5.137	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022519\
Data File : VX007702.D
Acq On : 26 Feb 2019 08:14
Operator : JC/SP
Sample : K1603-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
M2-T3-UV(PRE)

Quant Time: Feb 26 11:50:21 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 01:53:13 2019
Response via : Initial Calibration

