

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022519\
 Data File : VX007706.D
 Acq On : 26 Feb 2019 10:00
 Operator : JC/SP
 Sample : K1596-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COMP-13

Quant Time: Feb 26 12:01:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 01:53:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	326941	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	555487	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	546764	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	241302	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	199475	49.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.52%	
35) Dibromofluoromethane	5.51	113	186285	50.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.34%	
50) Toluene-d8	8.71	98	701014	49.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.62%	
62) 4-Bromofluorobenzene	11.13	95	243944	46.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.50%	
Target Compounds						
16) Acetone	2.45	43	19073	10.591	ug/l	99
18) Methyl Acetate	2.78	43	8420	2.143	ug/l	94
20) Methylene Chloride	2.86	84	106590	32.053	ug/l	96
43) Isopropyl Acetate	6.46	43	21011	2.461	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022519\
Data File : VX007706.D
Acq On : 26 Feb 2019 10:00
Operator : JC/SP
Sample : K1596-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
COMP-13

Quant Time: Feb 26 12:01:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 01:53:13 2019
Response via : Initial Calibration

