

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022521\
 Data File : VX020914.D
 Acq On : 25 Feb 2021 15:35
 Operator : JC/MD
 Sample : M1481-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-025-IDWGW-20210223

Manual Integrations
 APPROVED

MMDadoda
 2/26/2021 2:30:59 PM

Quant Time: Feb 25 16:10:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X022521W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Feb 25 13:26:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.988	128	36971m	30.00	ug/l	0.01
28) 1,4-Difluorobenzene	6.835	114	208748	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.094	117	175167	30.00	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	6.035	65	80193	30.80	ug/l	0.01
Spiked Amount	30.000	Range	91 - 110	Recovery	=	102.67%
60) 4-Bromofluorobenzene	11.118	95	81697	28.20	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	94.00%
63) Toluene-d8	8.694	98	242026	30.34	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	101.13%
Target Compounds						Qvalue
3) Chloromethane	1.317	50	7207	3.24	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022521\
Data File : VX020914.D
Acq On : 25 Feb 2021 15:35
Operator : JC/MD
Sample : M1481-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-025-IDWGW-20210223

Manual Integrations
APPROVED

MMDadoda
2/26/2021 2:30:59 PM

Quant Time: Feb 25 16:10:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X022521W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Feb 25 13:26:35 2021
Response via : Initial Calibration

