

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022521\
 Data File : VX020915.D
 Acq On : 25 Feb 2021 15:58
 Operator : JC/MD
 Sample : M1481-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-026-IDWGW-20210223

Quant Time: Feb 25 16:26:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X022521W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Feb 25 13:26:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.982	128	35958	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.829	114	196554	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.094	117	171564	30.00	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	6.029	65	76842	30.35	ug/l	0.00
Spiked Amount 30.000	Range	91 - 110	Recovery	=	101.17%	
60) 4-Bromofluorobenzene	11.118	95	79540	28.03	ug/l	0.00
Spiked Amount 30.000	Range	63 - 112	Recovery	=	93.43%	
63) Toluene-d8	8.694	98	239998	30.72	ug/l	0.00
Spiked Amount 30.000	Range	91 - 112	Recovery	=	102.40%	
Target Compounds						
22) cis-1,2-Dichloroethene	4.558	96	4041	1.50	ug/l	90
37) Trichloroethene	7.188	130	25971	10.37	ug/l	98
39) 1,2-Dichloropropane	7.488	63	2552	1.11	ug/l	92
61) Tetrachloroethene	9.318	164	1522	0.72	ug/l	91
91) Naphthalene	13.818	128	4507	0.46	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022521\
Data File : VX020915.D
Acq On : 25 Feb 2021 15:58
Operator : JC/MD
Sample : M1481-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-026-IDWGW-20210223

Quant Time: Feb 25 16:26:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X022521W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Feb 25 13:26:35 2021
Response via : Initial Calibration

