

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022521\
 Data File : VX020916.D
 Acq On : 25 Feb 2021 16:19
 Operator : JC/MD
 Sample : M1481-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-027-IDWGW-20210223

Quant Time: Feb 25 16:41:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X022521W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Feb 25 13:26:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.982	128	37746	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.829	114	212125	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.094	117	178079	30.00	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	6.035	65	80240	30.19	ug/l	0.01
Spiked Amount	30.000	Range	91 - 110	Recovery	=	100.63%
60) 4-Bromofluorobenzene	11.118	95	83484	28.34	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	94.47%
63) Toluene-d8	8.694	98	246067	30.35	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	101.17%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022521\
Data File : VX020916.D
Acq On : 25 Feb 2021 16:19
Operator : JC/MD
Sample : M1481-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-027-IDWGW-20210223

Quant Time: Feb 25 16:41:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X022521W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Feb 25 13:26:35 2021
Response via : Initial Calibration

