

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022522\
 Data File : VX027183.D
 Acq On : 25 Feb 2022 12:50
 Operator : JC/MD
 Sample : VX0225WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0225WBL01

Quant Time: Feb 25 23:16:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	96543	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	158821	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	144846	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	67585	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	57907	49.942	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.880%
35) Dibromofluoromethane	5.391	113	49303	50.012	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.020%
50) Toluene-d8	8.653	98	181724	49.299	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.600%
62) 4-Bromofluorobenzene	11.079	95	68840	48.653	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022522\
Data File : VX027183.D
Acq On : 25 Feb 2022 12:50
Operator : JC/MD
Sample : VX0225WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0225WBL01

Quant Time: Feb 25 23:16:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

