

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022522\
 Data File : VX027191.D
 Acq On : 25 Feb 2022 16:06
 Operator : JC/MD
 Sample : N1676-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Feb 25 23:18:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	91826	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	151406	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	136890	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	62392	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	55242	50.091	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.180%	
35) Dibromofluoromethane	5.397	113	47404	50.440	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.880%	
50) Toluene-d8	8.653	98	174400	49.630	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.260%	
62) 4-Bromofluorobenzene	11.079	95	65111	48.271	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.540%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022522\
Data File : VX027191.D
Acq On : 25 Feb 2022 16:06
Operator : JC/MD
Sample : N1676-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Feb 25 23:18:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

