

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022522\
 Data File : VX027192.D
 Acq On : 25 Feb 2022 16:30
 Operator : JC/MD
 Sample : N1676-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Feb 25 23:18:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	91069	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	151447	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	137429	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	62258	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	54543	49.868	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.740%
35) Dibromofluoromethane	5.391	113	46805	49.790	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.580%
50) Toluene-d8	8.653	98	175863	50.032	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.060%
62) 4-Bromofluorobenzene	11.085	95	67410	49.962	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022522\
Data File : VX027192.D
Acq On : 25 Feb 2022 16:30
Operator : JC/MD
Sample : N1676-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Feb 25 23:18:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

