

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022522\
 Data File : VX027194.D
 Acq On : 25 Feb 2022 17:17
 Operator : JC/MD
 Sample : N1676-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Feb 25 23:18:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	90087	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	149825	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	136794	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	62812	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	54336	50.220	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.440%	
35) Dibromofluoromethane	5.391	113	46609	50.118	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.240%	
50) Toluene-d8	8.653	98	171379	49.284	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.560%	
62) 4-Bromofluorobenzene	11.085	95	65412	49.006	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.020%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022522\
Data File : VX027194.D
Acq On : 25 Feb 2022 17:17
Operator : JC/MD
Sample : N1676-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Feb 25 23:18:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

