

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX022522\  
 Data File : VX027196.D  
 Acq On : 25 Feb 2022 18:03  
 Operator : JC/MD  
 Sample : PB142919ZHE#01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 17 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 PB142919ZHE#01

Quant Time: Feb 25 23:19:08 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X022322W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Feb 23 13:06:57 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units      | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|------------|----------|
| -----                       |        |       |          |          |            |          |
| Internal Standards          |        |       |          |          |            |          |
| 1) Pentafluorobenzene       | 5.556  | 168   | 88397    | 50.000   | ug/l       | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.769  | 114   | 145477   | 50.000   | ug/l       | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 132672   | 50.000   | ug/l       | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 59970    | 50.000   | ug/l       | 0.00     |
| System Monitoring Compounds |        |       |          |          |            |          |
| 33) 1,2-Dichloroethane-d4   | 5.964  | 65    | 53296    | 50.201   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | = 100.400% |          |
| 35) Dibromofluoromethane    | 5.391  | 113   | 45384    | 50.259   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | = 100.520% |          |
| 50) Toluene-d8              | 8.653  | 98    | 169983   | 50.344   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | = 100.680% |          |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 63323    | 48.859   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | = 97.720%  |          |
| Target Compounds            |        |       |          |          |            |          |
|                             |        |       |          |          | Qvalue     |          |
| 16) Acetone                 | 2.380  | 43    | 6441     | 13.818   | ug/l       | 100      |
| 20) Methylene Chloride      | 2.794  | 84    | 35816    | 37.072   | ug/l       | 94       |
| 43) Isopropyl Acetate       | 6.355  | 43    | 8321     | 3.738    | ug/l       | 99       |
| -----                       |        |       |          |          |            |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX022522\  
Data File : VX027196.D  
Acq On : 25 Feb 2022 18:03  
Operator : JC/MD  
Sample : PB142919ZHE#01  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 17 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
PB142919ZHE#01

Quant Time: Feb 25 23:19:08 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X022322W.M  
Quant Title : SW846 8260  
QLast Update : Wed Feb 23 13:06:57 2022  
Response via : Initial Calibration

