

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022522\
 Data File : VX027200.D
 Acq On : 25 Feb 2022 19:37
 Operator : JC/MD
 Sample : PB142919ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB142919ZHE#05

Quant Time: Feb 25 23:20:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	91512	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	150077	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	136372	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	60878	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	54459	49.550	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	99.100%		
35) Dibromofluoromethane	5.391	113	47702	51.207	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	102.420%		
50) Toluene-d8	8.653	98	177885	51.070	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	102.140%		
62) 4-Bromofluorobenzene	11.079	95	65168	48.741	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	97.480%		
Target Compounds						
16) Acetone	2.386	43	8830	18.298	ug/l	97
20) Methylene Chloride	2.794	84	58299	58.289	ug/l	95
43) Isopropyl Acetate	6.348	43	8271	3.602	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022522\
Data File : VX027200.D
Acq On : 25 Feb 2022 19:37
Operator : JC/MD
Sample : PB142919ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB142919ZHE#05

Quant Time: Feb 25 23:20:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

