

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022522\
 Data File : VX027208.D
 Acq On : 25 Feb 2022 23:31
 Operator : JC/MD
 Sample : VX0225MBL02
 Misc : 5.00g/10.0mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0225MBL02

Quant Time: Feb 26 00:39:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	95897	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	158192	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	143045	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	63414	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	58711	50.976	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 101.960%	
35) Dibromofluoromethane	5.391	113	49900	50.819	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.640%	
50) Toluene-d8	8.653	98	182102	49.598	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.200%	
62) 4-Bromofluorobenzene	11.079	95	66836	47.424	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 94.840%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022522\
Data File : VX027208.D
Acq On : 25 Feb 2022 23:31
Operator : JC/MD
Sample : VX0225MBL02
Misc : 5.00g/10.0mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0225MBL02

Quant Time: Feb 26 00:39:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

