

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022522\
 Data File : VX027212.D
 Acq On : 26 Feb 2022 01:04
 Operator : JC/MD
 Sample : N1670-01ME
 Misc : 5.09g/10.0mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OR-03-022422ME

Quant Time: Feb 26 02:48:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	90043	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	150914	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	138577	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	66875	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	54311	50.222	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.440%
35) Dibromofluoromethane	5.391	113	45168	48.218	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.440%
50) Toluene-d8	8.653	98	173741	49.603	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.200%
62) 4-Bromofluorobenzene	11.085	95	66793	49.679	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.360%
Target Compounds						
25) 2-Butanone	4.605	43	7552	9.987	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022522\
Data File : VX027212.D
Acq On : 26 Feb 2022 01:04
Operator : JC/MD
Sample : N1670-01ME
Misc : 5.09g/10.0mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OR-03-022422ME

Quant Time: Feb 26 02:48:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

