

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022522\
 Data File : VX027213.D
 Acq On : 26 Feb 2022 01:28
 Operator : JC/MD
 Sample : N1584-07ME
 Misc : 5.35g/5.0mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 D13-7.5ME

Quant Time: Feb 26 02:49:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	95792	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	160791	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	147042	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74750	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	57400	49.893	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	99.780%		
35) Dibromofluoromethane	5.397	113	46833	46.924	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.840%		
50) Toluene-d8	8.653	98	187557	50.258	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	100.520%		
62) 4-Bromofluorobenzene	11.085	95	73061	51.003	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	102.000%		
Target Compounds						Qvalue
31) Cyclohexane	5.495	56	3355	1.947	ug/l #	49
39) Methylcyclohexane	7.373	83	5340	2.947	ug/l	95
52) Toluene	8.720	92	9940	3.644	ug/l	100
67) Ethyl Benzene	10.195	91	22494	4.275	ug/l	98
68) m/p-Xylenes	10.305	106	25036	12.376	ug/l	97
69) o-Xylene	10.646	106	8129	4.004	ug/l	99
73) Isopropylbenzene	10.963	105	9569	1.789	ug/l	97
78) n-propylbenzene	11.305	91	24093	3.963	ug/l	98
80) 1,3,5-Trimethylbenzene	11.457	105	22060	4.905	ug/l	100
84) 1,2,4-Trimethylbenzene	11.756	105	57459	12.843	ug/l	97
89) n-Butylbenzene	12.335	91	11091	2.839	ug/l #	43
95) Naphthalene	13.780	128	10549	2.008	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022522\
Data File : VX027213.D
Acq On : 26 Feb 2022 01:28
Operator : JC/MD
Sample : N1584-07ME
Misc : 5.35g/5.0mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
D13-7.5ME

Quant Time: Feb 26 02:49:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

