

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030325\
 Data File : VX045100.D
 Acq On : 03 Mar 2025 08:47
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Mar 03 09:18:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	110336	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	6.751	114	188916	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	164466	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	76846	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	78535	44.75	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.50%
35) Dibromofluoromethane	5.373	113	60858	48.18	ug/l	-0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.36%
50) Toluene-d8	8.641	98	228601	49.92	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.84%
62) 4-Bromofluorobenzene	11.079	95	79497	52.38	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	104.76%

Target Compounds					Qvalue
2) Dichlorodifluoromethane	1.166	85	76304	54.94	ug/l 98
3) Chloromethane	1.294	50	91333	53.74	ug/l 98
4) Vinyl Chloride	1.374	62	86098	51.06	ug/l 96
5) Bromomethane	1.593	94	34492	52.04	ug/l 97
6) Chloroethane	1.672	64	48810	62.76	ug/l 97
7) Trichlorofluoromethane	1.880	101	119695	53.62	ug/l 98
8) Diethyl Ether	2.130	74	44057	50.46	ug/l 100
9) 1,1,2-Trichlorotrifluo...	2.325	101	71981	56.13	ug/l 99
10) Methyl Iodide	2.447	142	85221	53.18	ug/l 99
11) Tert butyl alcohol	2.965	59	63467m	190.92	ug/l
12) 1,1-Dichloroethene	2.312	96	69125	50.98	ug/l 97
13) Acrolein	2.233	56	83160	216.40	ug/l 99
14) Allyl chloride	2.660	41	127308	52.75	ug/l 97
15) Acrylonitrile	3.056	53	197379	221.53	ug/l 100
16) Acetone	2.373	43	202880	249.15	ug/l 97
17) Carbon Disulfide	2.508	76	194098	53.75	ug/l 97
18) Methyl Acetate	2.697	43	83486	42.62	ug/l 100
19) Methyl tert-butyl Ether	3.105	73	222567	48.93	ug/l 97
20) Methylene Chloride	2.782	84	76690	47.54	ug/l 98
21) trans-1,2-Dichloroethene	3.081	96	69898	52.18	ug/l 96
22) Diisopropyl ether	3.751	45	248151	50.56	ug/l 91
23) Vinyl Acetate	3.709	43	1056619	257.97	ug/l 99
24) 1,1-Dichloroethane	3.599	63	134862	49.03	ug/l 99
25) 2-Butanone	4.544	43	280054	228.48	ug/l 98
26) 2,2-Dichloropropane	4.465	77	104804	81.23	ug/l 95
27) cis-1,2-Dichloroethene	4.477	96	82263	49.75	ug/l 95
28) Bromochloromethane	4.885	49	66902	51.03	ug/l 99
29) Tetrahydrofuran	4.995	42	169123	205.46	ug/l 100
30) Chloroform	5.080	83	133151	48.72	ug/l 100
31) Cyclohexane	5.464	56	125240	52.45	ug/l 99
32) 1,1,1-Trichloroethane	5.373	97	111844	50.67	ug/l 100
36) 1,1-Dichloropropene	5.684	75	93232	53.82	ug/l 99
37) Ethyl Acetate	4.696	43	101438	45.44	ug/l 99
38) Carbon Tetrachloride	5.666	117	97961	55.70	ug/l 94
39) Methylcyclohexane	7.373	83	124590	60.02	ug/l 97
40) Benzene	6.025	78	283933	51.90	ug/l 98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030325\
 Data File : VX045100.D
 Acq On : 03 Mar 2025 08:47
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Mar 03 09:18:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.903	41	57735	47.83	ug/l	98
42) 1,2-Dichloroethane	6.074	62	98234	49.87	ug/l	100
43) Isopropyl Acetate	6.330	43	165728	50.21	ug/l	99
44) Trichloroethene	7.117	130	67789	52.77	ug/l	99
45) 1,2-Dichloropropane	7.421	63	71127	50.57	ug/l	99
46) Dibromomethane	7.574	93	51508	51.81	ug/l	99
47) Bromodichloromethane	7.811	83	103145	52.90	ug/l	99
48) Methyl methacrylate	7.683	41	81543	49.24	ug/l	99
49) 1,4-Dioxane	7.653	88	33724	965.40	ug/l	98
51) 4-Methyl-2-Pentanone	8.567	43	527249	237.89	ug/l	100
52) Toluene	8.714	92	171620	53.47	ug/l	100
53) t-1,3-Dichloropropene	8.970	75	99191	60.85	ug/l	96
54) cis-1,3-Dichloropropene	8.360	75	112314	59.08	ug/l	98
55) 1,1,2-Trichloroethane	9.147	97	65000	49.19	ug/l	96
56) Ethyl methacrylate	9.110	69	105447	52.50	ug/l	99
57) 1,3-Dichloropropane	9.305	76	114030	49.87	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.232	63	263987	269.44	ug/l	99
59) 2-Hexanone	9.427	43	396897	247.15	ug/l	100
60) Dibromochloromethane	9.512	129	75321	54.52	ug/l	99
61) 1,2-Dibromoethane	9.604	107	67274	51.00	ug/l	99
64) Tetrachloroethene	9.269	164	56666	53.80	ug/l	96
65) Chlorobenzene	10.073	112	183860	52.83	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.159	131	61384	52.97	ug/l	98
67) Ethyl Benzene	10.189	91	328535	54.18	ug/l	99
68) m/p-Xylenes	10.299	106	243777	109.82	ug/l	99
69) o-Xylene	10.640	106	118940	53.13	ug/l	100
70) Styrene	10.652	104	202784	55.98	ug/l	100
71) Bromoform	10.799	173	49189	56.28	ug/l #	98
73) Isopropylbenzene	10.957	105	311255	51.89	ug/l	100
74) N-amyl acetate	10.841	43	147759	52.50	ug/l	100
75) 1,1,2,2-Tetrachloroethane	11.207	83	103046	46.82	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	108619	60.73	ug/l	86
77) Bromobenzene	11.195	156	73426	51.88	ug/l	99
78) n-propylbenzene	11.299	91	369377	54.52	ug/l	100
79) 2-Chlorotoluene	11.360	91	218391	50.61	ug/l	99
80) 1,3,5-Trimethylbenzene	11.445	105	258263	53.25	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.012	75	29745	56.91	ug/l	99
82) 4-Chlorotoluene	11.451	91	249399	53.03	ug/l	99
83) tert-Butylbenzene	11.713	119	261634	52.50	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	261885	53.67	ug/l	100
85) sec-Butylbenzene	11.884	105	328305	55.00	ug/l	100
86) p-Isopropyltoluene	12.006	119	274625	56.93	ug/l	99
87) 1,3-Dichlorobenzene	11.963	146	134914	53.44	ug/l	100
88) 1,4-Dichlorobenzene	12.036	146	135078	52.89	ug/l	99
89) n-Butylbenzene	12.329	91	253261	60.96	ug/l	100
90) Hexachloroethane	12.536	117	47961	54.97	ug/l	97
91) 1,2-Dichlorobenzene	12.329	146	134345	53.05	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	20011	46.67	ug/l	100
93) 1,2,4-Trichlorobenzene	13.585	180	83758	58.08	ug/l	99
94) Hexachlorobutadiene	13.719	225	34572	58.48	ug/l	99
95) Naphthalene	13.774	128	280556	50.64	ug/l	99
96) 1,2,3-Trichlorobenzene	13.957	180	81985	53.89	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030325\
Data File : VX045100.D
Acq On : 03 Mar 2025 08:47
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Mar 03 09:18:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030325\
 Data File : VX045100.D
 Acq On : 03 Mar 2025 08:47
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :


Quant Time: Mar 03 09:18:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 5.538 min Scan# 71
Delta R.T. -0.012 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:168 Resp: 110336
Ion Ratio Lower Upper
168 100
99 60.9 48.2 72.4



#2
Dichlorodifluoromethane
Concen: 54.94 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Tgt Ion: 85 Resp: 76304
Ion Ratio Lower Upper
85 100
87 33.5 17.4 52.3









#3
Chloromethane
Concen: 53.74 ug/l
RT: 1.294 min Scan# 34
Delta R.T. -0.012 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 50 Resp: 91333
Ion Ratio Lower Upper
50 100
52 32.3 25.0 37.4



#4
Vinyl Chloride
Concen: 51.06 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47


Tgt Ion: 62 Resp: 86098
Ion Ratio Lower Upper
62 100
64 32.7 24.3 36.5








#5
Bromomethane
Concen: 52.04 ug/l
RT: 1.593 min Scan# 81
Delta R.T. -0.000 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 94 Resp: 34492
Ion Ratio Lower Upper
94 100
96 97.1 75.0 112.4



#6
Chloroethane
Concen: 62.76 ug/l
RT: 1.672 min Scan# 96
Delta R.T. 0.012 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Tgt Ion: 64 Resp: 48810
Ion Ratio Lower Upper
64 100
66 31.7 26.7 40.1









#7
Trichlorofluoromethane
Concen: 53.62 ug/l
RT: 1.880 min Scan# 11
Delta R.T. 0.012 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 119695
Ion Ratio Lower Upper
101 100
103 67.0 52.1 78.1



#8
Diethyl Ether
Concen: 50.46 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47


Tgt Ion: 74 Resp: 44057
Ion Ratio Lower Upper
74 100
45 102.8 51.5 154.5








#9
1,1,2-Trichlorotrifluoroethane
Concen: 56.13 ug/l
RT: 2.325 min Scan# 201
Delta R.T. 0.012 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:101 Resp: 71981
Ion Ratio Lower Upper
101 100
85 46.0 36.2 54.4
151 71.0 56.4 84.6



#10
Methyl Iodide
Concen: 53.18 ug/l
RT: 2.447 min Scan# 223
Delta R.T. 0.006 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Tgt Ion:142 Resp: 85221
Ion Ratio Lower Upper
142 100
127 45.2 35.4 53.2
141 14.5 11.6 17.4





#11

Tert butyl alcohol

Concen: 190.92 ug/l m

RT: 2.965 min Scan# 308

Delta R.T. -0.018 min

Lab File: VX045100.D

Acq: 03 Mar 2025 08:47

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 59 Resp: 63467

Ion Ratio Lower Upper

59 100

57 10.4 7.8 11.8



#12

1,1-Dichloroethene

Concen: 50.98 ug/l

RT: 2.312 min Scan# 201

Delta R.T. 0.006 min

Lab File: VX045100.D

Acq: 03 Mar 2025 08:47




Tgt Ion: 96 Resp: 69125


Ion Ratio Lower Upper

96 100

61 172.6 134.6 202.0


98 66.8 51.0 76.6








#13
Acrolein
Concen: 216.40 ug/l
RT: 2.233 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 56 Resp: 83160
Ion Ratio Lower Upper
56 100
55 70.8 56.2 84.4



#14
Allyl chloride
Concen: 52.75 ug/l
RT: 2.660 min Scan# 258
Delta R.T. 0.006 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47


Tgt Ion: 41 Resp: 127308
Ion Ratio Lower Upper
41 100
39 68.9 53.8 80.8
76 33.5 25.2 37.8








#15
Acrylonitrile
Concen: 221.53 ug/l
RT: 3.056 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 53 Resp: 197379
Ion Ratio Lower Upper
53 100
52 83.0 66.2 99.2
51 35.8 29.0 43.4



#16
Acetone
Concen: 249.15 ug/l
RT: 2.373 min Scan# 211
Delta R.T. -0.012 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Tgt Ion: 43 Resp: 202880
Ion Ratio Lower Upper
43 100
58 31.9 24.2 36.4









#17
Carbon Disulfide
Concen: 53.75 ug/l
RT: 2.508 min Scan# 21
Delta R.T. 0.006 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 76 Resp: 194098
Ion Ratio Lower Upper
76 100
78 9.3 6.6 9.8



#18
Methyl Acetate
Concen: 42.62 ug/l
RT: 2.697 min Scan# 264
Delta R.T. -0.006 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Tgt Ion: 43 Resp: 83486
Ion Ratio Lower Upper
43 100
74 22.0 17.4 26.2









#19
Methyl tert-butyl Ether
Concen: 48.93 ug/l
RT: 3.105 min Scan# 311
Delta R.T. -0.012 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 73 Resp: 222567
Ion Ratio Lower Upper
73 100
57 21.6 18.5 27.7



#20
Methylene Chloride
Concen: 47.54 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.000 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47


Tgt Ion: 84 Resp: 76690
Ion Ratio Lower Upper
84 100
49 133.7 106.5 159.7
51 42.4 32.1 48.1
86 65.2 50.0 75.0








#21
trans-1,2-Dichloroethene
Concen: 52.18 ug/l
RT: 3.081 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 96 Resp: 69898
Ion Ratio Lower Upper
96 100
61 142.7 117.0 175.4
98 61.6 53.4 80.2



#22
Diisopropyl ether
Concen: 50.56 ug/l
RT: 3.751 min Scan# 437
Delta R.T. -0.012 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Tgt Ion: 45 Resp: 248151
Ion Ratio Lower Upper
45 100
43 79.6 54.9 82.3
87 26.1 21.0 31.4
59 11.4 8.7 13.1









#23
Vinyl Acetate
Concen: 257.97 ug/l
RT: 3.709 min Scan# 411
Delta R.T. -0.012 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 1056619
Ion Ratio Lower Upper
43 100
86 9.7 8.1 12.1



#24
1,1-Dichloroethane
Concen: 49.03 ug/l
RT: 3.599 min Scan# 412
Delta R.T. -0.006 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Tgt Ion: 63 Resp: 134862
Ion Ratio Lower Upper
63 100
98 7.5 3.4 10.2
100 4.4 2.1 6.5









#25
2-Butanone
Concen: 228.48 ug/l
RT: 4.544 min Scan# 51
Delta R.T. -0.018 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 280054
Ion Ratio Lower Upper
43 100
72 23.8 20.0 30.0



#26
2,2-Dichloropropane
Concen: 81.23 ug/l
RT: 4.465 min Scan# 554
Delta R.T. -0.006 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Tgt Ion: 77 Resp: 104804
Ion Ratio Lower Upper
77 100
97 22.4 12.4 37.0









#27
 cis-1,2-Dichloroethene
 Concen: 49.75 ug/l
 RT: 4.477 min Scan# 51
 Delta R.T. -0.006 min
 Lab File: VX045100.D
 Acq: 03 Mar 2025 08:47




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 96 Resp: 82263
 Ion Ratio Lower Upper
 96 100
 61 149.7 0.0 283.2
 98 63.8 0.0 128.0



#28
 Bromochloromethane
 Concen: 51.03 ug/l
 RT: 4.885 min Scan# 623
 Delta R.T. -0.012 min
 Lab File: VX045100.D
 Acq: 03 Mar 2025 08:47

Tgt Ion: 49 Resp: 66902
 Ion Ratio Lower Upper
 49 100
 129 1.9 0.0 3.4
 130 69.5 56.1 84.1





#29

Tetrahydrofuran

Concen: 205.46 ug/l

RT: 4.995 min Scan# 641

Delta R.T. -0.012 min

Lab File: VX045100.D

Acq: 03 Mar 2025 08:47

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 42 Resp: 169123

Ion Ratio Lower Upper

42 100

72 42.7 34.1 51.1

71 38.8 31.4 47.0



#30

Chloroform

Concen: 48.72 ug/l

RT: 5.080 min Scan# 655

Delta R.T. -0.006 min

Lab File: VX045100.D




Acq: 03 Mar 2025 08:47


Tgt Ion: 83 Resp: 133151

Ion Ratio Lower Upper

83 100


85 64.0 51.4 77.2








#31
Cyclohexane
Concen: 52.45 ug/l
RT: 5.464 min Scan# 718
Delta R.T. -0.000 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 56 Resp: 125240
Ion Ratio Lower Upper
56 100
69 30.6 23.4 35.2
84 86.1 69.4 104.2



#32
1,1,1-Trichloroethane
Concen: 50.67 ug/l
RT: 5.373 min Scan# 703
Delta R.T. -0.006 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Tgt Ion: 97 Resp: 111844
Ion Ratio Lower Upper
97 100
99 64.9 51.6 77.4
61 47.8 38.4 57.6









#33
1,2-Dichloroethane-d4
Concen: 44.75 ug/l
RT: 5.946 min Scan# 798
Delta R.T. -0.006 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 65 Resp: 78535
Ion Ratio Lower Upper
65 100
67 53.1 0.0 106.2



#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 6.751 min Scan# 929
Delta R.T. -0.006 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47


Tgt Ion:114 Resp: 188916
Ion Ratio Lower Upper
114 100
63 22.0 0.0 41.8
88 15.8 0.0 32.8








#35
Dibromofluoromethane
Concen: 48.18 ug/l
RT: 5.373 min Scan# 705
Delta R.T. -0.012 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:113 Resp: 60858
Ion Ratio Lower Upper
113 100
111 104.6 81.8 122.6
192 18.7 14.3 21.5



#36
1,1-Dichloropropene
Concen: 53.82 ug/l
RT: 5.684 min Scan# 754
Delta R.T. -0.006 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47


Tgt Ion: 75 Resp: 93232
Ion Ratio Lower Upper
75 100
110 34.1 16.9 50.6
77 30.8 24.5 36.7








#37
Ethyl Acetate
Concen: 45.44 ug/l
RT: 4.696 min Scan# 51
Delta R.T. -0.018 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 43 Resp: 101438
Ion Ratio Lower Upper
43 100
61 13.3 10.8 16.2
70 9.3 7.7 11.5



#38
Carbon Tetrachloride
Concen: 55.70 ug/l
RT: 5.666 min Scan# 751
Delta R.T. -0.006 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Tgt Ion: 117 Resp: 97961
Ion Ratio Lower Upper
117 100
119 90.4 76.7 115.1
121 28.6 25.5 38.3





#39

Methylcyclohexane

Concen: 60.02 ug/l

RT: 7.373 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX045100.D

Acq: 03 Mar 2025 08:47

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 83 Resp: 124590

Ion Ratio Lower Upper

83 100

55 76.0 63.0 94.4

98 48.4 39.7 59.5



#40

Benzene

Concen: 51.90 ug/l

RT: 6.025 min Scan# 810

Delta R.T. -0.006 min

Lab File: VX045100.D




Acq: 03 Mar 2025 08:47


Tgt Ion: 78 Resp: 283933

Ion Ratio Lower Upper

78 100


77 24.4 18.8 28.2








#41
Methacrylonitrile
Concen: 47.83 ug/l
RT: 4.903 min Scan# 61
Delta R.T. -0.018 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion: 41 Resp: 57735




Ion	Ratio	Lower	Upper
41	100		
39	53.3	44.6	67.0
67	69.3	53.9	80.9
52	29.4	23.3	34.9




#42
1,2-Dichloroethane
Concen: 49.87 ug/l
RT: 6.074 min Scan# 818
Delta R.T. -0.006 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Tgt Ion: 62 Resp: 98234


Ion	Ratio	Lower	Upper
62	100		
98	9.0	0.0	18.2




Abundance Scan 862 (6.342 min): VX045071.D\data.ms (-85



m/z--> Scan 860 (6.330 min): VX045100.D\data.ms




Abundance Scan 860 (6.330 min): VX045100.D\data.ms (-81




m/z-->


Abundance Scan 990 (7.123 min): VX045071.D\data.ms (-98



m/z--> Scan 989 (7.117 min): VX045100.D\data.ms



Abundance Scan 989 (7.117 min): VX045100.D\data.ms (-94



m/z-->

#43

Isopropyl Acetate

Concen: 50.21 ug/l

RT: 6.330 min Scan# 860

Delta R.T. -0.012 min

Lab File: VX045100.D

Acq: 03 Mar 2025 08:47

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion: 43 Resp: 165728


Ion Ratio Lower Upper

43 100

61 18.4 14.9 22.3

87 12.4 9.4 14.2

Abundance



#44

Trichloroethene

Concen: 52.77 ug/l

RT: 7.117 min Scan# 989

Delta R.T. -0.006 min

Lab File: VX045100.D

Acq: 03 Mar 2025 08:47


Tgt Ion:130 Resp: 67789


Ion Ratio Lower Upper

130 100

95 101.8 0.0 205.0

Abundance





#45

1,2-Dichloropropane

Concen: 50.57 ug/l

RT: 7.421 min Scan# 1039

Delta R.T. -0.006 min

Lab File: VX045100.D

Acq: 03 Mar 2025 08:47

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 63 Resp: 71127

Ion Ratio Lower Upper

63 100

65 31.7 24.7 37.1



#46

Dibromomethane

Concen: 51.81 ug/l

RT: 7.574 min Scan# 1064

Delta R.T. -0.006 min

Lab File: VX045100.D

Acq: 03 Mar 2025 08:47




Tgt Ion: 93 Resp: 51508


Ion Ratio Lower Upper

93 100

95 83.9 65.8 98.8

174 89.1 72.2 108.2





#47

Bromodichloromethane

Concen: 52.90 ug/l

RT: 7.811 min Scan# 1103

Delta R.T. -0.006 min

Lab File: VX045100.D





Acq: 03 Mar 2025 08:47

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion:	83	Resp:	103145
Ion	Ratio	Lower	Upper
83	100		
85	63.1	51.1	76.7
127	7.8	6.4	9.6



#48

Methyl methacrylate

Concen: 49.24 ug/l




RT: 7.683 min Scan# 1082


Delta R.T. -0.006 min

Lab File: VX045100.D

Acq: 03 Mar 2025 08:47

Tgt Ion:	41	Resp:	81543
Ion	Ratio	Lower	Upper
41	100		
69	79.2	63.0	94.6
39	60.5	47.5	71.3





#49

1,4-Dioxane

Concen: 965.40 ug/l

RT: 7.653 min Scan# 1

Delta R.T. -0.006 min

Lab File: VX045100.D

Acq: 03 Mar 2025 08:47

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 88 Resp: 33724

Ion Ratio Lower Upper

88 100

43 35.9 28.7 43.1

58 67.9 55.8 83.8



#50

Toluene-d8

Concen: 49.92 ug/l

RT: 8.641 min Scan# 1239

Delta R.T. -0.006 min

Lab File: VX045100.D




Acq: 03 Mar 2025 08:47


Tgt Ion: 98 Resp: 228601

Ion Ratio Lower Upper

98 100

100 64.5 52.0 78.0





#51

4-Methyl-2-Pentanone

Concen: 237.89 ug/l

RT: 8.567 min Scan# 11

Delta R.T. -0.006 min

Lab File: VX045100.D

Acq: 03 Mar 2025 08:47

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 43 Resp: 527249

Ion Ratio Lower Upper

43 100

58 36.8 29.2 43.8



#52

Toluene

Concen: 53.47 ug/l

RT: 8.714 min Scan# 1251

Delta R.T. -0.000 min

Lab File: VX045100.D




Acq: 03 Mar 2025 08:47


Tgt Ion: 92 Resp: 171620

Ion Ratio Lower Upper

92 100

91 173.2 138.9 208.3





#53

t-1,3-Dichloropropene

Concen: 60.85 ug/l

RT: 8.970 min Scan# 1193

Delta R.T. -0.006 min

Lab File: VX045100.D

Acq: 03 Mar 2025 08:47

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 75 Resp: 99191

Ion Ratio Lower Upper

75 100

77 31.0 26.5 39.7



#54

cis-1,3-Dichloropropene

Concen: 59.08 ug/l

RT: 8.360 min Scan# 1193

Delta R.T. -0.006 min

Lab File: VX045100.D

Acq: 03 Mar 2025 08:47




Tgt Ion: 75 Resp: 112314


Ion Ratio Lower Upper

75 100

77 31.0 24.7 37.1

39 53.4 40.7 61.1









#55
1,1,2-Trichloroethane
Concen: 49.19 ug/l
RT: 9.147 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 97 Resp: 65000




Ion	Ratio	Lower	Upper
97	100		
83	89.7	69.4	104.2
85	59.3	44.1	66.1
99	65.3	50.7	76.1




#56
Ethyl methacrylate
Concen: 52.50 ug/l
RT: 9.110 min Scan# 1316
Delta R.T. -0.006 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Tgt Ion: 69 Resp: 105447

Ion	Ratio	Lower	Upper
69	100		
41	72.7	57.0	85.4
39	42.6	34.2	51.4









#57
1,3-Dichloropropane
Concen: 49.87 ug/l
RT: 9.305 min Scan# 114030
Delta R.T. -0.000 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 76 Resp: 114030
Ion Ratio Lower Upper
76 100
78 32.5 25.5 38.3



#58
2-Chloroethyl Vinyl ether
Concen: 269.44 ug/l
RT: 8.232 min Scan# 1172
Delta R.T. -0.006 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Tgt Ion: 63 Resp: 263987
Ion Ratio Lower Upper
63 100
106 26.4 21.5 32.3





#59

2-Hexanone

Concen: 247.15 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. -0.000 min

Lab File: VX045100.D

Acq: 03 Mar 2025 08:47

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 43 Resp: 396897

Ion Ratio Lower Upper

43 100

58 51.6 25.9 77.6



#60

Dibromochloromethane

Concen: 54.52 ug/l

RT: 9.512 min Scan# 1382

Delta R.T. -0.006 min

Lab File: VX045100.D




Acq: 03 Mar 2025 08:47


Tgt Ion:129 Resp: 75321

Ion Ratio Lower Upper

129 100

127 77.8 38.5 115.5









#61
1,2-Dibromoethane
Concen: 51.00 ug/l
RT: 9.604 min Scan# 1397
Delta R.T. -0.000 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:107 Resp: 67274
Ion Ratio Lower Upper
107 100
109 93.8 76.1 114.1



#62
4-Bromofluorobenzene
Concen: 52.38 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47


Tgt Ion: 95 Resp: 79497
Ion Ratio Lower Upper
95 100
174 74.5 0.0 148.2
176 72.8 0.0 141.4








#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:117 Resp: 164466
Ion Ratio Lower Upper
117 100
82 56.3 46.3 69.5
119 32.1 25.7 38.5



#64
Tetrachloroethene
Concen: 53.80 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Tgt Ion:164 Resp: 56666
Ion Ratio Lower Upper
164 100
166 130.7 103.6 155.4
129 97.1 82.7 124.1
131 95.9 80.5 120.7









#65
Chlorobenzene
Concen: 52.83 ug/l
RT: 10.073 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:112 Resp: 183860
Ion Ratio Lower Upper
112 100
114 31.5 26.0 39.0



#66
1,1,1,2-Tetrachloroethane
Concen: 52.97 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Tgt Ion:131 Resp: 61384
Ion Ratio Lower Upper
131 100
133 94.7 48.6 145.9
119 67.3 33.9 101.7







#67
Ethyl Benzene
Concen: 54.18 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. -0.000 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47


Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 328535
Ion Ratio Lower Upper
91 100
106 29.6 23.9 35.9



#68
m/p-Xylenes
Concen: 109.82 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Tgt Ion:106 Resp: 243777
Ion Ratio Lower Upper
106 100
91 208.2 165.4 248.0









#69
o-Xylene
Concen: 53.13 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion:106 Resp: 118940
Ion Ratio Lower Upper
106 100
91 219.3 109.9 329.6




#70
Styrene
Concen: 55.98 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Tgt Ion:104 Resp: 202784
Ion Ratio Lower Upper
104 100
78 52.6 42.2 63.4
103 54.9 43.8 65.8



Abundance Scan 1593 (10.799 min): VX045071.D\data.ms (-



#71

Bromoform

Concen: 56.28 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. -0.000 min

Lab File: VX045100.D

Acq: 03 Mar 2025 08:47

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion:173 Resp: 49189


Ion Ratio Lower Upper

173 100


175 47.9 24.6 73.6

254 0.0 0.0 0.0


Abundance Scan 1593 (10.799 min): VX045100.D\data.ms




Abundance



Abundance Scan 1593 (10.799 min): VX045100.D\data.ms (-



Abundance Scan 1793 (12.018 min): VX045071.D\data.ms (-



#72

1,4-Dichlorobenzene-d4

Concen: 50.00 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. -0.000 min

Lab File: VX045100.D

Acq: 03 Mar 2025 08:47

Tgt Ion:152 Resp: 76846


Ion Ratio Lower Upper

152 100


115 89.3 44.2 132.6

150 178.3 0.0 349.0


Abundance Scan 1793 (12.018 min): VX045100.D\data.ms




Abundance



Abundance Scan 1793 (12.018 min): VX045100.D\data.ms (-









#73
Isopropylbenzene
Concen: 51.89 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.006 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:105 Resp: 311255
Ion Ratio Lower Upper
105 100
120 26.4 13.2 39.5



#74
N-ethyl acetate
Concen: 52.50 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Tgt Ion: 43 Resp: 147759
Ion Ratio Lower Upper
43 100
70 39.3 31.8 47.6
55 22.5 18.3 27.5
61 22.9 18.3 27.5









#75
1,1,2,2-Tetrachloroethane
Concen: 46.82 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. -0.006 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 103046
Ion Ratio Lower Upper
83 100
131 9.3 4.5 13.4
85 64.3 31.7 95.1



#76
1,2,3-Trichloropropane
Concen: 60.73 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Tgt Ion: 75 Resp: 108619
Ion Ratio Lower Upper
75 100
77 32.8 20.7 62.1









#77
Bromobenzene
Concen: 51.88 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:156 Resp: 73426
Ion Ratio Lower Upper
156 100
77 169.5 85.8 257.4
158 95.5 48.4 145.2



#78
n-propylbenzene
Concen: 54.52 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. -0.006 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Tgt Ion: 91 Resp: 369377
Ion Ratio Lower Upper
91 100
120 22.6 11.2 33.6









#79
2-Chlorotoluene
Concen: 50.61 ug/l
RT: 11.360 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 218391
Ion Ratio Lower Upper
91 100
126 32.5 16.4 49.4



#80
1,3,5-Trimethylbenzene
Concen: 53.25 ug/l
RT: 11.445 min Scan# 1699
Delta R.T. -0.006 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47


Tgt Ion: 105 Resp: 258263
Ion Ratio Lower Upper
105 100
120 47.8 24.1 72.2








#81
trans-1,4-Dichloro-2-butene
Concen: 56.91 ug/l
RT: 11.012 min Scan# 1629
Delta R.T. -0.006 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 75 Resp: 29745
Ion Ratio Lower Upper
75 100
53 109.4 88.5 132.7
89 44.5 36.2 54.4




#82
4-Chlorotoluene
Concen: 53.03 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47



Tgt Ion: 91 Resp: 249399
Ion Ratio Lower Upper
91 100
126 28.6 14.1 42.4









#83
tert-Butylbenzene
Concen: 52.50 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. -0.000 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:119 Resp: 261634
Ion Ratio Lower Upper
119 100
91 59.7 30.7 92.1
134 22.9 11.7 35.0



#84
1,2,4-Trimethylbenzene
Concen: 53.67 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Tgt Ion:105 Resp: 261885
Ion Ratio Lower Upper
105 100
120 43.9 22.1 66.1









#85
sec-Butylbenzene
Concen: 55.00 ug/l
RT: 11.884 min Scan# 1772
Delta R.T. -0.006 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:105 Resp: 328305
Ion Ratio Lower Upper
105 100
134 19.4 9.8 29.4



#86
p-Isopropyltoluene
Concen: 56.93 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47


Tgt Ion:119 Resp: 274625
Ion Ratio Lower Upper
119 100
134 25.2 12.9 38.6
91 24.4 12.7 38.0








#87
1,3-Dichlorobenzene
Concen: 53.44 ug/l
RT: 11.963 min Scan# 1784
Delta R.T. -0.006 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:146 Resp: 134914
Ion Ratio Lower Upper
146 100
111 42.2 21.1 63.4
148 63.6 31.9 95.7



#88
1,4-Dichlorobenzene
Concen: 52.89 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. -0.000 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47


Tgt Ion:146 Resp: 135078
Ion Ratio Lower Upper
146 100
111 40.9 20.2 60.5
148 63.4 31.9 95.9








#89
n-Butylbenzene
Concen: 60.96 ug/l
RT: 12.329 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 91 Resp: 253261
Ion Ratio Lower Upper
91 100
92 53.4 26.7 80.0
134 23.9 12.2 36.6



#90
Hexachloroethane
Concen: 54.97 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. -0.000 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Tgt Ion: 117 Resp: 47961
Ion Ratio Lower Upper
117 100
201 66.2 31.9 95.9









#91
1,2-Dichlorobenzene
Concen: 53.05 ug/l
RT: 12.329 min Scan# 1844
Delta R.T. -0.006 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:146 Resp: 134345
Ion Ratio Lower Upper
146 100
111 42.7 21.6 65.0
148 62.8 31.9 95.9



#92
1,2-Dibromo-3-Chloropropane
Concen: 46.67 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47


Tgt Ion: 75 Resp: 20011
Ion Ratio Lower Upper
75 100
155 78.4 39.6 118.7
157 102.3 51.1 153.4








#93
1,2,4-Trichlorobenzene
Concen: 58.08 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. -0.000 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:180 Resp: 83758
Ion Ratio Lower Upper
180 100
182 94.1 47.3 141.9
145 31.8 16.3 48.9



#94
Hexachlorobutadiene
Concen: 58.48 ug/l
RT: 13.719 min Scan# 2072
Delta R.T. -0.006 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47


Tgt Ion:225 Resp: 34572
Ion Ratio Lower Upper
225 100
223 63.5 31.9 95.7
227 63.2 31.4 94.0








#95
Naphthalene
Concen: 50.64 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion:128 Resp: 280556
Ion Ratio Lower Upper
128 100
127 12.8 10.3 15.5
129 11.2 8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 53.89 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX045100.D
Acq: 03 Mar 2025 08:47

Tgt Ion:180 Resp: 81985
Ion Ratio Lower Upper
180 100
182 96.7 47.5 142.6
145 34.5 17.0 50.9

