

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030525\
 Data File : VX045127.D
 Acq On : 05 Mar 2025 09:12
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Mar 05 09:42:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	112291	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	190470	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	164207	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	74481	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	5.946	65	83482	46.74	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.48%
35) Dibromofluoromethane	5.373	113	65550	51.47	ug/l	-0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.94%
50) Toluene-d8	8.641	98	234119	50.70	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.40%
62) 4-Bromofluorobenzene	11.079	95	78409	51.25	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	102.50%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.166	85	68622	48.55	ug/l	91
3) Chloromethane	1.307	50	75368	43.57	ug/l	97
4) Vinyl Chloride	1.374	62	77400	45.11	ug/l	98
5) Bromomethane	1.593	94	30416	45.09	ug/l	98
6) Chloroethane	1.660	64	32683	41.29	ug/l	95
7) Trichlorofluoromethane	1.867	101	105280	46.35	ug/l	98
8) Diethyl Ether	2.136	74	39098	44.00	ug/l	97
9) 1,1,2-Trichlorotrifluo...	2.312	101	65196	49.95	ug/l	99
10) Methyl Iodide	2.441	142	78420	48.08	ug/l	98
11) Tert butyl alcohol	2.995	59	65619m	193.95	ug/l	
12) 1,1-Dichloroethene	2.306	96	62054	44.97	ug/l	99
13) Acrolein	2.233	56	93541	239.18	ug/l	98
14) Allyl chloride	2.654	41	116846	47.57	ug/l	98
15) Acrylonitrile	3.062	53	196064	216.23	ug/l	99
16) Acetone	2.386	43	168809	203.70	ug/l	97
17) Carbon Disulfide	2.495	76	163042	44.36	ug/l	97
18) Methyl Acetate	2.703	43	91008	45.65	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	208272	44.99	ug/l	100
20) Methylene Chloride	2.782	84	71831	43.76	ug/l	97
21) trans-1,2-Dichloroethene	3.081	96	64065	47.00	ug/l	97
22) Diisopropyl ether	3.757	45	232828	46.61	ug/l	96
23) Vinyl Acetate	3.715	43	988338	237.10	ug/l	99
24) 1,1-Dichloroethane	3.599	63	126104	45.05	ug/l	98
25) 2-Butanone	4.556	43	274151	219.77	ug/l	97
26) 2,2-Dichloropropane	4.465	77	89223	67.95	ug/l	96
27) cis-1,2-Dichloroethene	4.477	96	76838	45.66	ug/l	95
28) Bromochloromethane	4.891	49	62721	47.00	ug/l	99
29) Tetrahydrofuran	5.001	42	179664	214.47	ug/l	100
30) Chloroform	5.086	83	126092	45.33	ug/l	98
31) Cyclohexane	5.458	56	116269	47.85	ug/l	99
32) 1,1,1-Trichloroethane	5.373	97	103689	46.15	ug/l	99
36) 1,1-Dichloropropene	5.678	75	86057	49.27	ug/l	99
37) Ethyl Acetate	4.715	43	105492	46.87	ug/l	99
38) Carbon Tetrachloride	5.666	117	87042	49.09	ug/l	99
39) Methylcyclohexane	7.373	83	115913	55.38	ug/l	98
40) Benzene	6.025	78	269708	48.90	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030525\
 Data File : VX045127.D
 Acq On : 05 Mar 2025 09:12
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Mar 05 09:42:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	59812	49.15	ug/l	98
42) 1,2-Dichloroethane	6.080	62	96363	48.52	ug/l	100
43) Isopropyl Acetate	6.336	43	161345	48.48	ug/l	100
44) Trichloroethene	7.117	130	64443	49.76	ug/l	99
45) 1,2-Dichloropropane	7.421	63	67543	47.63	ug/l	97
46) Dibromomethane	7.574	93	48542	48.43	ug/l	98
47) Bromodichloromethane	7.818	83	97568	49.63	ug/l	100
48) Methyl methacrylate	7.690	41	80625	48.29	ug/l	99
49) 1,4-Dioxane	7.665	88	31601	897.25	ug/l	97
51) 4-Methyl-2-Pentanone	8.574	43	523954	234.47	ug/l	99
52) Toluene	8.714	92	163330	50.47	ug/l	98
53) t-1,3-Dichloropropene	8.976	75	88322	53.74	ug/l	99
54) cis-1,3-Dichloropropene	8.360	75	102986	53.73	ug/l	97
55) 1,1,2-Trichloroethane	9.147	97	63538	47.69	ug/l	96
56) Ethyl methacrylate	9.116	69	98764	48.77	ug/l	99
57) 1,3-Dichloropropane	9.305	76	110247	47.82	ug/l	98
58) 2-Chloroethyl Vinyl ether	8.238	63	263234	266.48	ug/l	100
59) 2-Hexanone	9.427	43	384958	237.76	ug/l	100
60) Dibromochloromethane	9.518	129	71321	51.20	ug/l	98
61) 1,2-Dibromoethane	9.604	107	64667	48.62	ug/l	99
64) Tetrachloroethene	9.269	164	53835	51.19	ug/l	97
65) Chlorobenzene	10.073	112	171449	49.34	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	56543	48.87	ug/l	99
67) Ethyl Benzene	10.189	91	308282	50.92	ug/l	99
68) m/p-Xylenes	10.299	106	228130	102.93	ug/l	100
69) o-Xylene	10.640	106	111076	49.69	ug/l	99
70) Styrene	10.652	104	183603	50.76	ug/l	100
71) Bromoform	10.799	173	44654	51.18	ug/l #	99
73) Isopropylbenzene	10.957	105	289986	49.88	ug/l	100
74) N-amyl acetate	10.841	43	133706	49.01	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	97087	45.51	ug/l	98
76) 1,2,3-Trichloropropane	11.238	75	100313	57.87	ug/l	87
77) Bromobenzene	11.195	156	65952	48.08	ug/l	99
78) n-propylbenzene	11.305	91	338544	51.55	ug/l	99
79) 2-Chlorotoluene	11.360	91	196917	47.08	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	235799	50.16	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	26035	51.40	ug/l	98
82) 4-Chlorotoluene	11.451	91	223976	49.14	ug/l	100
83) tert-Butylbenzene	11.713	119	232845	48.21	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	236246	49.95	ug/l	100
85) sec-Butylbenzene	11.890	105	298121	51.53	ug/l	99
86) p-Isopropyltoluene	12.006	119	242004	51.76	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	119022	48.64	ug/l	100
88) 1,4-Dichlorobenzene	12.042	146	118036	47.68	ug/l	99
89) n-Butylbenzene	12.329	91	222008	55.13	ug/l	100
90) Hexachloroethane	12.536	117	42059	49.73	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	121609	49.54	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	19675	47.34	ug/l	98
93) 1,2,4-Trichlorobenzene	13.585	180	73807	52.81	ug/l	100
94) Hexachlorobutadiene	13.725	225	30136	52.59	ug/l	97
95) Naphthalene	13.774	128	269735	50.23	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	75650	51.31	ug/l	100


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030525\
Data File : VX045127.D
Acq On : 05 Mar 2025 09:12
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Mar 05 09:42:22 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion:168 Resp: 112291
Ion Ratio Lower Upper
168 100
99 61.8 48.2 72.4






#2
Dichlorodifluoromethane
Concen: 48.55 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Tgt Ion: 85 Resp: 68622
Ion Ratio Lower Upper
85 100
87 29.8 17.4 52.3



Abundance Scan 36 (1.307 min): VX045071.D\data.ms (-29)



#3

Chloromethane

Concen: 43.57 ug/l

RT: 1.307 min Scan# 36

Delta R.T. -0.000 min

Lab File: VX045127.D

Acq: 05 Mar 2025 09:12

Instrument :

MSVOA_X


ClientSampleId :

Tgt Ion: 50 Resp: 75368




Ion Ratio Lower Upper

50 100

52 33.1 25.0 37.4



Abundance Scan 47 (1.374 min): VX045071.D\data.ms (-41)



#4

Vinyl Chloride

Concen: 45.11 ug/l

RT: 1.374 min Scan# 47

Delta R.T. -0.000 min

Lab File: VX045127.D


Acq: 05 Mar 2025 09:12


Tgt Ion: 62 Resp: 77400

Ion Ratio Lower Upper

62 100


64 31.3 24.3 36.5








#5
Bromomethane
Concen: 45.09 ug/l
RT: 1.593 min Scan# 81
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Instrument :
MSVOA_X
ClientSampleId :






Tgt Ion: 94 Resp: 30416
Ion Ratio Lower Upper
94 100
96 91.6 75.0 112.4




#6
Chloroethane
Concen: 41.29 ug/l
RT: 1.660 min Scan# 94
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Tgt Ion: 64 Resp: 32683
Ion Ratio Lower Upper
64 100
66 30.3 26.7 40.1



Abundance Scan 128 (1.868 min): VX045071.D\data.ms (-12



#7

Trichlorofluoromethane

Concen: 46.35 ug/l

RT: 1.867 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX045127.D

Acq: 05 Mar 2025 09:12

Instrument :

MSVOA_X

ClientSampleId :


Tgt Ion:101 Resp: 105280

Ion Ratio Lower Upper


101 100

103 66.5 52.1 78.1


Abundance Scan 128 (1.867 min): VX045127.D\data.ms




Abundance



Abundance Scan 128 (1.867 min): VX045127.D\data.ms (-79



Abundance Scan 172 (2.136 min): VX045071.D\data.ms (-16



#8

Diethyl Ether

Concen: 44.00 ug/l

RT: 2.136 min Scan# 172

Delta R.T. -0.000 min

Lab File: VX045127.D

Acq: 05 Mar 2025 09:12


Tgt Ion: 74 Resp: 39098

Ion Ratio Lower Upper


74 100

45 106.0 51.5 154.5


Abundance Scan 172 (2.136 min): VX045127.D\data.ms




Abundance



Abundance Scan 172 (2.136 min): VX045127.D\data.ms (-12









#9
1,1,2-Trichlorotrifluoroethane
Concen: 49.95 ug/l
RT: 2.312 min Scan# 201
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 65196
Ion Ratio Lower Upper
101 100
85 46.1 36.2 54.4
151 71.3 56.4 84.6



#10
Methyl Iodide
Concen: 48.08 ug/l
RT: 2.441 min Scan# 222
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Tgt Ion:142 Resp: 78420
Ion Ratio Lower Upper
142 100
127 45.8 35.4 53.2
141 14.2 11.6 17.4





#11

Tert butyl alcohol

Concen: 193.95 ug/l m

RT: 2.995 min Scan# 311

Delta R.T. 0.012 min

Lab File: VX045127.D

Acq: 05 Mar 2025 09:12

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 59 Resp: 65619

Ion Ratio Lower Upper

59 100

57 9.2 7.8 11.8



#12

1,1-Dichloroethene

Concen: 44.97 ug/l

RT: 2.306 min Scan# 200

Delta R.T. -0.000 min

Lab File: VX045127.D

Acq: 05 Mar 2025 09:12




Tgt Ion: 96 Resp: 62054


Ion Ratio Lower Upper

96 100

61 169.5 134.6 202.0


98 62.8 51.0 76.6








#13
Acrolein
Concen: 239.18 ug/l
RT: 2.233 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 56 Resp: 93541
Ion Ratio Lower Upper
56 100
55 71.8 56.2 84.4



#14
Allyl chloride
Concen: 47.57 ug/l
RT: 2.654 min Scan# 257
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Tgt Ion: 41 Resp: 116846
Ion Ratio Lower Upper
41 100
39 68.4 53.8 80.8
76 33.0 25.2 37.8









#15
Acrylonitrile
Concen: 216.23 ug/l
RT: 3.062 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 53 Resp: 196064
Ion Ratio Lower Upper
53 100
52 83.4 66.2 99.2
51 35.9 29.0 43.4



#16
Acetone
Concen: 203.70 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Tgt Ion: 43 Resp: 168809
Ion Ratio Lower Upper
43 100
58 28.7 24.2 36.4









#17
Carbon Disulfide
Concen: 44.36 ug/l
RT: 2.495 min Scan# 21
Delta R.T. -0.006 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 76 Resp: 163042
Ion Ratio Lower Upper
76 100
78 9.4 6.6 9.8



#18
Methyl Acetate
Concen: 45.65 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Tgt Ion: 43 Resp: 91008
Ion Ratio Lower Upper
43 100
74 22.2 17.4 26.2









#19
Methyl tert-butyl Ether
Concen: 44.99 ug/l
RT: 3.111 min Scan# 311
Delta R.T. -0.006 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 73 Resp: 208272
Ion Ratio Lower Upper
73 100
57 23.1 18.5 27.7



#20
Methylene Chloride
Concen: 43.76 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12


Tgt Ion: 84 Resp: 71831
Ion Ratio Lower Upper
84 100
49 128.6 106.5 159.7
51 37.9 32.1 48.1
86 63.2 50.0 75.0








#21
trans-1,2-Dichloroethene
Concen: 47.00 ug/l
RT: 3.081 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 96 Resp: 64065
Ion Ratio Lower Upper
96 100
61 143.1 117.0 175.4
98 64.6 53.4 80.2



#22
Diisopropyl ether
Concen: 46.61 ug/l
RT: 3.757 min Scan# 438
Delta R.T. -0.006 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Tgt Ion: 45 Resp: 232828
Ion Ratio Lower Upper
45 100
43 73.2 54.9 82.3
87 25.8 21.0 31.4
59 10.6 8.7 13.1









#23
 Vinyl Acetate
 Concen: 237.10 ug/l
 RT: 3.715 min Scan# 41
 Delta R.T. -0.006 min
 Lab File: VX045127.D
 Acq: 05 Mar 2025 09:12




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 43 Resp: 988338
 Ion Ratio Lower Upper
 43 100
 86 9.9 8.1 12.1



#24
 1,1-Dichloroethane
 Concen: 45.05 ug/l
 RT: 3.599 min Scan# 412
 Delta R.T. -0.006 min
 Lab File: VX045127.D
 Acq: 05 Mar 2025 09:12

Tgt Ion: 63 Resp: 126104
 Ion Ratio Lower Upper
 63 100
 98 6.4 3.4 10.2
 100 4.9 2.1 6.5





#25

2-Butanone

Concen: 219.77 ug/l

RT: 4.556 min Scan# 5

Delta R.T. -0.006 min

Lab File: VX045127.D

Acq: 05 Mar 2025 09:12

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 43 Resp: 274151

Ion Ratio Lower Upper

43 100

72 23.6 20.0 30.0



#26

2,2-Dichloropropane

Concen: 67.95 ug/l

RT: 4.465 min Scan# 554

Delta R.T. -0.006 min

Lab File: VX045127.D




Acq: 05 Mar 2025 09:12


Tgt Ion: 77 Resp: 89223

Ion Ratio Lower Upper

77 100


97 22.8 12.4 37.0








#27
cis-1,2-Dichloroethene
Concen: 45.66 ug/l
RT: 4.477 min Scan# 51
Delta R.T. -0.006 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 96 Resp: 76838
Ion Ratio Lower Upper
96 100
61 150.1 0.0 283.2
98 64.4 0.0 128.0



#28
Bromochloromethane
Concen: 47.00 ug/l
RT: 4.891 min Scan# 624
Delta R.T. -0.006 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Tgt Ion: 49 Resp: 62721
Ion Ratio Lower Upper
49 100
129 1.9 0.0 3.4
130 70.7 56.1 84.1









#29
Tetrahydrofuran
Concen: 214.47 ug/l
RT: 5.001 min Scan# 643
Delta R.T. -0.006 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 42 Resp: 179664
Ion Ratio Lower Upper
42 100
72 42.4 34.1 51.1
71 39.1 31.4 47.0



#30
Chloroform
Concen: 45.33 ug/l
RT: 5.086 min Scan# 656
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12


Tgt Ion: 83 Resp: 126092
Ion Ratio Lower Upper
83 100
85 65.8 51.4 77.2








#31
Cyclohexane
Concen: 47.85 ug/l
RT: 5.458 min Scan# 718
Delta R.T. -0.006 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 56 Resp: 116269
Ion Ratio Lower Upper
56 100
69 29.4 23.4 35.2
84 85.8 69.4 104.2



#32
1,1,1-Trichloroethane
Concen: 46.15 ug/l
RT: 5.373 min Scan# 703
Delta R.T. -0.006 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Tgt Ion: 97 Resp: 103689
Ion Ratio Lower Upper
97 100
99 64.5 51.6 77.4
61 49.2 38.4 57.6









#33
1,2-Dichloroethane-d4
Concen: 46.74 ug/l
RT: 5.946 min Scan# 798
Delta R.T. -0.006 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 65 Resp: 83482
Ion Ratio Lower Upper
65 100
67 52.5 0.0 106.2



#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 6.751 min Scan# 929
Delta R.T. -0.006 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12


Tgt Ion: 114 Resp: 190470
Ion Ratio Lower Upper
114 100
63 22.2 0.0 41.8
88 16.4 0.0 32.8








#35
Dibromofluoromethane
Concen: 51.47 ug/l
RT: 5.373 min Scan# 705
Delta R.T. -0.012 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Instrument :
MSVOA_X
ClientSampleId :






Tgt Ion:113 Resp: 65550
Ion Ratio Lower Upper
113 100
111 102.1 81.8 122.6
192 17.7 14.3 21.5






#36
1,1-Dichloropropene
Concen: 49.27 ug/l
RT: 5.678 min Scan# 753
Delta R.T. -0.012 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Tgt Ion: 75 Resp: 86057
Ion Ratio Lower Upper
75 100
110 35.0 16.9 50.6
77 31.0 24.5 36.7



Abundance Scan 595 (4.715 min): VX045071.D\data.ms (-58



#37

Ethyl Acetate

Concen: 46.87 ug/l

RT: 4.715 min Scan# 595

Delta R.T. -0.000 min

Lab File: VX045127.D

Acq: 05 Mar 2025 09:12

Instrument :

MSVOA_X

ClientSampleId :


Tgt Ion: 43 Resp: 105492

Ion Ratio Lower Upper




43 100

61 13.2 10.8 16.2

70 9.4 7.7 11.5



Abundance Scan 752 (5.672 min): VX045071.D\data.ms (-74



#38

Carbon Tetrachloride

Concen: 49.09 ug/l

RT: 5.666 min Scan# 751

Delta R.T. -0.006 min

Lab File: VX045127.D

Acq: 05 Mar 2025 09:12


Tgt Ion: 117 Resp: 87042


Ion Ratio Lower Upper

117 100

119 95.3 76.7 115.1

121 30.7 25.5 38.3





#39

Methylcyclohexane

Concen: 55.38 ug/l

RT: 7.373 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX045127.D

Acq: 05 Mar 2025 09:12

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 83 Resp: 115913

Ion Ratio Lower Upper

83 100

55 76.4 63.0 94.4

98 48.3 39.7 59.5



#40

Benzene

Concen: 48.90 ug/l

RT: 6.025 min Scan# 810

Delta R.T. -0.006 min

Lab File: VX045127.D




Acq: 05 Mar 2025 09:12


Tgt Ion: 78 Resp: 269708

Ion Ratio Lower Upper

78 100


77 23.5 18.8 28.2








#41
Methacrylonitrile
Concen: 49.15 ug/l
RT: 4.916 min Scan# 61
Delta R.T. -0.006 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion: 41 Resp: 59812




Ion	Ratio	Lower	Upper
41	100		
39	53.4	44.6	67.0
67	67.2	53.9	80.9
52	30.6	23.3	34.9




#42
1,2-Dichloroethane
Concen: 48.52 ug/l
RT: 6.080 min Scan# 819
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Tgt Ion: 62 Resp: 96363


Ion	Ratio	Lower	Upper
62	100		
98	9.1	0.0	18.2




Abundance Scan 862 (6.342 min): VX045071.D\data.ms (-85



m/z--> Scan 861 (6.336 min): VX045127.D\data.ms



Abundance Scan 861 (6.336 min): VX045127.D\data.ms (-81



m/z-->

#43

Isopropyl Acetate

Concen: 48.48 ug/l

RT: 6.336 min Scan# 861

Delta R.T. -0.006 min

Lab File: VX045127.D

Acq: 05 Mar 2025 09:12

Instrument :

MSVOA_X

ClientSampleId :


Tgt Ion: 43 Resp: 161345

Ion Ratio Lower Upper


43 100

61 18.8 14.9 22.3


87 12.0 9.4 14.2




Abundance Scan 990 (7.123 min): VX045071.D\data.ms (-98



m/z--> Scan 989 (7.117 min): VX045127.D\data.ms



Abundance Scan 989 (7.117 min): VX045127.D\data.ms (-94



m/z-->

#44

Trichloroethene

Concen: 49.76 ug/l

RT: 7.117 min Scan# 989

Delta R.T. -0.006 min

Lab File: VX045127.D


Acq: 05 Mar 2025 09:12


Tgt Ion:130 Resp: 64443

Ion Ratio Lower Upper

130 100

95 104.0 0.0 205.0









#45
1,2-Dichloropropane
Concen: 47.63 ug/l
RT: 7.421 min Scan# 1039
Delta R.T. -0.006 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 63 Resp: 67543
Ion Ratio Lower Upper
63 100
65 32.6 24.7 37.1



#46
Dibromomethane
Concen: 48.43 ug/l
RT: 7.574 min Scan# 1064
Delta R.T. -0.006 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Tgt Ion: 93 Resp: 48542
Ion Ratio Lower Upper
93 100
95 85.0 65.8 98.8
174 89.7 72.2 108.2









#47
Bromodichloromethane
Concen: 49.63 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 97568
Ion Ratio Lower Upper
83 100
85 64.3 51.1 76.7
127 8.1 6.4 9.6



#48
Methyl methacrylate
Concen: 48.29 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12


Tgt Ion: 41 Resp: 80625
Ion Ratio Lower Upper
41 100
69 78.0 63.0 94.6
39 60.5 47.5 71.3








#49
1,4-Dioxane
Concen: 897.25 ug/l
RT: 7.665 min Scan# 1078
Delta R.T. 0.006 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 88 Resp: 31601
Ion Ratio Lower Upper
88 100
43 36.7 28.7 43.1
58 72.4 55.8 83.8



#50
Toluene-d8
Concen: 50.70 ug/l
RT: 8.641 min Scan# 1239
Delta R.T. -0.006 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Tgt Ion: 98 Resp: 234119
Ion Ratio Lower Upper
98 100
100 66.2 52.0 78.0









#51
4-Methyl-2-Pentanone
Concen: 234.47 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 523954
Ion Ratio Lower Upper
43 100
58 36.9 29.2 43.8



#52
Toluene
Concen: 50.47 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Tgt Ion: 92 Resp: 163330
Ion Ratio Lower Upper
92 100
91 171.0 138.9 208.3









#53
t-1,3-Dichloropropene
Concen: 53.74 ug/l
RT: 8.976 min Scan# 1193
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 88322
Ion Ratio Lower Upper
75 100
77 32.5 26.5 39.7



#54
cis-1,3-Dichloropropene
Concen: 53.73 ug/l
RT: 8.360 min Scan# 1193
Delta R.T. -0.006 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Tgt Ion: 75 Resp: 102986
Ion Ratio Lower Upper
75 100
77 31.3 24.7 37.1
39 54.1 40.7 61.1









#55
1,1,2-Trichloroethane
Concen: 47.69 ug/l
RT: 9.147 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 97 Resp: 63538




Ion	Ratio	Lower	Upper
97	100		
83	91.5	69.4	104.2
85	58.4	44.1	66.1
99	63.8	50.7	76.1




#56
Ethyl methacrylate
Concen: 48.77 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Tgt Ion: 69 Resp: 98764

Ion	Ratio	Lower	Upper
69	100		
41	72.0	57.0	85.4
39	44.0	34.2	51.4









#57
1,3-Dichloropropane
Concen: 47.82 ug/l
RT: 9.305 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 76 Resp: 110247
Ion Ratio Lower Upper
76 100
78 32.8 25.5 38.3



#58
2-Chloroethyl Vinyl ether
Concen: 266.48 ug/l
RT: 8.238 min Scan# 1173
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Tgt Ion: 63 Resp: 263234
Ion Ratio Lower Upper
63 100
106 26.9 21.5 32.3









#59
2-Hexanone
Concen: 237.76 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 384958
Ion Ratio Lower Upper
43 100
58 52.0 25.9 77.6



#60
Dibromochloromethane
Concen: 51.20 ug/l
RT: 9.518 min Scan# 1383
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Tgt Ion: 129 Resp: 71321
Ion Ratio Lower Upper
129 100
127 75.5 38.5 115.5









#61
1,2-Dibromoethane
Concen: 48.62 ug/l
RT: 9.604 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:107 Resp: 64667
Ion Ratio Lower Upper
107 100
109 94.3 76.1 114.1



#62
4-Bromofluorobenzene
Concen: 51.25 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12


Tgt Ion: 95 Resp: 78409
Ion Ratio Lower Upper
95 100
174 74.7 0.0 148.2
176 71.1 0.0 141.4








#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:117 Resp: 164207
Ion Ratio Lower Upper
117 100
82 59.1 46.3 69.5
119 32.3 25.7 38.5



#64
Tetrachloroethene
Concen: 51.19 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Tgt Ion:164 Resp: 53835
Ion Ratio Lower Upper
164 100
166 129.5 103.6 155.4
129 98.2 82.7 124.1
131 97.4 80.5 120.7









#65
Chlorobenzene
Concen: 49.34 ug/l
RT: 10.073 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:112 Resp: 171449
Ion Ratio Lower Upper
112 100
114 32.2 26.0 39.0



#66
1,1,1,2-Tetrachloroethane
Concen: 48.87 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Tgt Ion:131 Resp: 56543
Ion Ratio Lower Upper
131 100
133 95.7 48.6 145.9
119 67.0 33.9 101.7





#67

Ethyl Benzene

Concen: 50.92 ug/l

RT: 10.189 min Scan# 1493

Delta R.T. -0.000 min

Lab File: VX045127.D

Acq: 05 Mar 2025 09:12

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 91 Resp: 308282

Ion Ratio Lower Upper

91 100

106 29.4 23.9 35.9



#68

m/p-Xylenes

Concen: 102.93 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX045127.D




Acq: 05 Mar 2025 09:12


Tgt Ion:106 Resp: 228130

Ion Ratio Lower Upper

106 100


91 206.7 165.4 248.0








#69
o-Xylene
Concen: 49.69 ug/l
RT: 10.640 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:106 Resp: 111076
Ion Ratio Lower Upper
106 100
91 218.6 109.9 329.6



#70
Styrene
Concen: 50.76 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12


Tgt Ion:104 Resp: 183603
Ion Ratio Lower Upper
104 100
78 52.8 42.2 63.4
103 54.9 43.8 65.8








#71
Bromoform
Concen: 51.18 ug/l
RT: 10.799 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:173 Resp: 44654
Ion Ratio Lower Upper
173 100
175 48.1 24.6 73.6
254 0.0 0.0 0.0




#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12



Tgt Ion:152 Resp: 74481
Ion Ratio Lower Upper
152 100
115 85.2 44.2 132.6
150 170.7 0.0 349.0









#73
Isopropylbenzene
Concen: 49.88 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.006 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:105 Resp: 289986
Ion Ratio Lower Upper
105 100
120 26.1 13.2 39.5



#74
N-ethyl acetate
Concen: 49.01 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Tgt Ion: 43 Resp: 133706
Ion Ratio Lower Upper
43 100
70 39.3 31.8 47.6
55 22.9 18.3 27.5
61 22.5 18.3 27.5









#75
1,1,2,2-Tetrachloroethane
Concen: 45.51 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 97087
Ion Ratio Lower Upper
83 100
131 9.2 4.5 13.4
85 64.8 31.7 95.1



#76
1,2,3-Trichloropropane
Concen: 57.87 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12


Tgt Ion: 75 Resp: 100313
Ion Ratio Lower Upper
75 100
77 33.2 20.7 62.1








#77
Bromobenzene
Concen: 48.08 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:156 Resp: 65952
Ion Ratio Lower Upper
156 100
77 169.8 85.8 257.4
158 97.1 48.4 145.2



#78
n-propylbenzene
Concen: 51.55 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Tgt Ion: 91 Resp: 338544
Ion Ratio Lower Upper
91 100
120 22.0 11.2 33.6









#79
2-Chlorotoluene
Concen: 47.08 ug/l
RT: 11.360 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 196917
Ion Ratio Lower Upper
91 100
126 33.0 16.4 49.4



#80
1,3,5-Trimethylbenzene
Concen: 50.16 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12


Tgt Ion:105 Resp: 235799
Ion Ratio Lower Upper
105 100
120 47.2 24.1 72.2








#81
trans-1,4-Dichloro-2-butene
Concen: 51.40 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 75 Resp: 26035
Ion Ratio Lower Upper
75 100
53 113.7 88.5 132.7
89 45.5 36.2 54.4



#82
4-Chlorotoluene
Concen: 49.14 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12


Tgt Ion: 91 Resp: 223976
Ion Ratio Lower Upper
91 100
126 28.4 14.1 42.4








#83
tert-Butylbenzene
Concen: 48.21 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:119 Resp: 232845
Ion Ratio Lower Upper
119 100
91 61.5 30.7 92.1
134 23.3 11.7 35.0



#84
1,2,4-Trimethylbenzene
Concen: 49.95 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Tgt Ion:105 Resp: 236246
Ion Ratio Lower Upper
105 100
120 44.0 22.1 66.1









#85
 sec-Butylbenzene
 Concen: 51.53 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX045127.D
 Acq: 05 Mar 2025 09:12




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion:105 Resp: 298121
 Ion Ratio Lower Upper
 105 100
 134 19.2 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 51.76 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.000 min
 Lab File: VX045127.D
 Acq: 05 Mar 2025 09:12


Tgt Ion:119 Resp: 242004
 Ion Ratio Lower Upper
 119 100
 134 26.1 12.9 38.6
 91 25.0 12.7 38.0








#87
1,3-Dichlorobenzene
Concen: 48.64 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:146 Resp: 119022
Ion Ratio Lower Upper
146 100
111 42.5 21.1 63.4
148 63.8 31.9 95.7



#88
1,4-Dichlorobenzene
Concen: 47.68 ug/l
RT: 12.042 min Scan# 1797
Delta R.T. 0.006 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Tgt Ion:146 Resp: 118036
Ion Ratio Lower Upper
146 100
111 40.9 20.2 60.5
148 64.7 31.9 95.9









#89
n-Butylbenzene
Concen: 55.13 ug/l
RT: 12.329 min Scan# 1844
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 222008
Ion Ratio Lower Upper
91 100
92 53.5 26.7 80.0
134 24.5 12.2 36.6



#90
Hexachloroethane
Concen: 49.73 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12


Tgt Ion: 117 Resp: 42059
Ion Ratio Lower Upper
117 100
201 65.0 31.9 95.9








#91
1,2-Dichlorobenzene
Concen: 49.54 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:146 Resp: 121609
Ion Ratio Lower Upper
146 100
111 42.6 21.6 65.0
148 62.7 31.9 95.9



#92
1,2-Dibromo-3-Chloropropane
Concen: 47.34 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12


Tgt Ion: 75 Resp: 19675
Ion Ratio Lower Upper
75 100
155 76.3 39.6 118.7
157 101.6 51.1 153.4








#93
1,2,4-Trichlorobenzene
Concen: 52.81 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:180 Resp: 73807
Ion Ratio Lower Upper
180 100
182 94.8 47.3 141.9
145 32.9 16.3 48.9




#94
Hexachlorobutadiene
Concen: 52.59 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12




Tgt Ion:225 Resp: 30136
Ion Ratio Lower Upper
225 100
223 62.9 31.9 95.7
227 65.7 31.4 94.0








#95
Naphthalene
Concen: 50.23 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion:128 Resp: 269735
Ion Ratio Lower Upper
128 100
127 13.0 10.3 15.5
129 10.8 8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 51.31 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.006 min
Lab File: VX045127.D
Acq: 05 Mar 2025 09:12

Tgt Ion:180 Resp: 75650
Ion Ratio Lower Upper
180 100
182 95.5 47.5 142.6
145 34.3 17.0 50.9

