

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022525\
 Data File : VX045030.D
 Acq On : 25 Feb 2025 11:44
 Operator : JC/MD
 Sample : VX0225MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0225MBL01

Quant Time: Feb 25 12:10:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	76885	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	151709	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	139213	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	59840	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	60431	53.69	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 107.38%		
35) Dibromofluoromethane	5.379	113	51079	51.79	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 103.58%		
50) Toluene-d8	8.647	98	186639	50.04	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 100.08%		
62) 4-Bromofluorobenzene	11.079	95	64630	51.40	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 102.80%		
Target Compounds						
20) Methylene Chloride	2.782	84	309	0.28	ug/l	Qvalue # 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022525\
Data File : VX045030.D
Acq On : 25 Feb 2025 11:44
Operator : JC/MD
Sample : VX0225MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0225MBL01

Quant Time: Feb 25 12:10:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

