

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022525\
 Data File : VX045035.D
 Acq On : 25 Feb 2025 13:44
 Operator : JC/MD
 Sample : Q1401-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-192-TB-20250217

Quant Time: Feb 26 03:15:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	80365	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	162183	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	145520	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	60661	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	61600	52.363	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.720%
35) Dibromofluoromethane	5.385	113	50460	47.854	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.700%
50) Toluene-d8	8.647	98	186526	46.780	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.560%
62) 4-Bromofluorobenzene	11.079	95	63053	46.907	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	93.820%
Target Compounds						
16) Acetone	2.380	43	1123	2.375	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022525\
Data File : VX045035.D
Acq On : 25 Feb 2025 13:44
Operator : JC/MD
Sample : Q1401-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-192-TB-20250217

Quant Time: Feb 26 03:15:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

