

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022525\
 Data File : VX045037.D
 Acq On : 25 Feb 2025 14:30
 Operator : JC/MD
 Sample : Q1423-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-192-EB-20250221

Quant Time: Feb 26 03:16:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	74974	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	149361	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	135427	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	54882	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	60976	55.560	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.120%
35) Dibromofluoromethane	5.385	113	50727	52.237	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.480%
50) Toluene-d8	8.647	98	184947	50.365	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.740%
62) 4-Bromofluorobenzene	11.079	95	60645	48.988	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.980%
Target Compounds						
16) Acetone	2.379	43	7128	16.156	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022525\
Data File : VX045037.D
Acq On : 25 Feb 2025 14:30
Operator : JC/MD
Sample : Q1423-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-192-EB-20250221

Quant Time: Feb 26 03:16:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

