

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022525\
 Data File : VX045039.D
 Acq On : 25 Feb 2025 15:17
 Operator : JC/MD
 Sample : Q1401-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-192-GW-840-842

Quant Time: Feb 26 03:17:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 11 03:41:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	82149	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	161424	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	145575	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	60559	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63496	52.803	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 105.600%			
35) Dibromofluoromethane	5.385	113	53442	50.920	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.840%			
50) Toluene-d8	8.647	98	190401	47.976	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 95.960%			
62) 4-Bromofluorobenzene	11.079	95	63599	47.536	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 95.080%			
Target Compounds					Qvalue	
16) Acetone	2.380	43	9621	19.902	ug/l	89
25) 2-Butanone	4.586	43	3931	5.008	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX022525\
Data File : VX045039.D
Acq On : 25 Feb 2025 15:17
Operator : JC/MD
Sample : Q1401-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-192-GW-840-842

Quant Time: Feb 26 03:17:20 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021025W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 11 03:41:08 2025
Response via : Initial Calibration

