

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022619\
 Data File : VX007712.D
 Acq On : 26 Feb 2019 14:39
 Operator : JC/SP
 Sample : K1578-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB201-EB-20190221

Quant Time: Feb 26 15:37:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 01:53:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	348276	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	594755	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	559199	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	253758	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	211698	49.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.14%	
35) Dibromofluoromethane	5.50	113	197090	49.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.14%	
50) Toluene-d8	8.71	98	724363	47.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.18%	
62) 4-Bromofluorobenzene	11.13	95	250869	44.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.82%	
Target Compounds						
16) Acetone	2.45	43	13832	7.210	ug/l	93
52) Toluene	8.79	92	6579	0.632	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022619\
Data File : VX007712.D
Acq On : 26 Feb 2019 14:39
Operator : JC/SP
Sample : K1578-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB201-EB-20190221

Quant Time: Feb 26 15:37:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 01:53:13 2019
Response via : Initial Calibration

