

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022619\
 Data File : VX007714.D
 Acq On : 26 Feb 2019 15:33
 Operator : JC/SP
 Sample : K1578-13DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB201-GW-340-342DL

Quant Time: Feb 26 16:04:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 01:53:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	327829	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	571193	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	529960	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	236603	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	208606	51.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.74%	
35) Dibromofluoromethane	5.50	113	190699	49.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.88%	
50) Toluene-d8	8.71	98	683778	46.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.56%	
62) 4-Bromofluorobenzene	11.13	95	237015	44.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.36%	
Target Compounds						
7) Trichlorofluoromethane	1.94	101	23177	4.403	ug/l	98
9) 1,1,2-Trichlorotrifluoroet	2.39	101	3421	1.075	ug/l #	83
12) 1,1-Dichloroethene	2.38	96	8858	3.089	ug/l	98
27) cis-1,2-Dichloroethene	4.60	96	3885	0.979	ug/l	86
44) Trichloroethene	7.21	130	189752	43.514	ug/l	92
64) Tetrachloroethene	9.34	164	89588	20.823	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022619\
Data File : VX007714.D
Acq On : 26 Feb 2019 15:33
Operator : JC/SP
Sample : K1578-13DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB201-GW-340-342DL

Quant Time: Feb 26 16:04:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 01:53:13 2019
Response via : Initial Calibration

