

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022619\
 Data File : VX007723.D
 Acq On : 26 Feb 2019 19:32
 Operator : JC/SP
 Sample : K1631-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z31W

Quant Time: Feb 27 00:49:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 01:53:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	337375	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	577203	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	541310	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	245124	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	213192	51.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.02%	
35) Dibromofluoromethane	5.51	113	193986	50.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.56%	
50) Toluene-d8	8.71	98	706330	47.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.64%	
62) 4-Bromofluorobenzene	11.13	95	239820	44.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.46%	
Target Compounds						
16) Acetone	2.45	43	4906	2.640	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022619\
Data File : VX007723.D
Acq On : 26 Feb 2019 19:32
Operator : JC/SP
Sample : K1631-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z31W

Quant Time: Feb 27 00:49:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 01:53:13 2019
Response via : Initial Calibration

