

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022619\
 Data File : VX007727.D
 Acq On : 26 Feb 2019 21:19
 Operator : JC/SP
 Sample : K1610-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WV-B

Quant Time: Feb 27 01:02:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 01:53:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	331173	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	576808	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	563150	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	249175	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.07	65	213084	51.94	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	103.88%	
35) Dibromofluoromethane	5.51	113	190761	49.47	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	98.94%	
50) Toluene-d8	8.71	98	722005	48.91	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	97.82%	
62) 4-Bromofluorobenzene	11.13	95	249771	46.10	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	92.20%	

Target Compounds

						Qvalue
16) Acetone	2.45	43	21895	12.002	ug/l #	88
18) Methyl Acetate	2.78	43	14449	3.630	ug/l	97
20) Methylene Chloride	2.86	84	393201	116.729	ug/l	95
43) Isopropyl Acetate	6.46	43	17747	2.002	ug/l #	90
95) Naphthalene	13.83	128	26702	1.629	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022619\
 Data File : VX007727.D
 Acq On : 26 Feb 2019 21:19
 Operator : JC/SP
 Sample : K1610-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 Client Sampled :
 WV-B

Quant Time: Feb 27 01:02:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 01:53:13 2019
 Response via : Initial Calibration

