

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX022620\
 Data File : VX015023.D
 Acq On : 26 Feb 2020 16:07
 Operator : JC/SP
 Sample : L1683-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Feb 27 02:56:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 10 12:09:20 2020
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	416676	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	661407	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	643700	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	329674	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	238866	50.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.60%	
35) Dibromofluoromethane	5.49	113	204489	50.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.58%	
50) Toluene-d8	8.71	98	797358	51.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.74%	
62) 4-Bromofluorobenzene	11.13	95	293861	52.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.12%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.19	85	1458	0.369	ug/l	82
3) Chloromethane	1.32	50	1591	0.356	ug/l #	54
4) Vinyl Chloride	1.40	62	1721	0.357	ug/l	93
5) Bromomethane	1.65	94	776	0.249	ug/l	93
9) 1,1,2-Trichlorotrifluoroet	2.39	101	1737	0.444	ug/l #	65
10) Methyl Iodide	2.51	142	252	4.421	ug/l #	1
11) Tert butyl alcohol	3.04	59	411	0.495	ug/l #	20
12) 1,1-Dichloroethene	2.37	96	4859	1.272	ug/l	97
13) Acrolein	2.28	56	388	0.613	ug/l #	15
16) Acetone	2.44	43	3563	Below Cal		96
21) trans-1,2-Dichloroethene	3.16	96	6084	1.444	ug/l	86
25) 2-Butanone	4.65	43	714	0.223	ug/l #	65
27) cis-1,2-Dichloroethene	4.59	96	177581	37.696	ug/l	89
28) Bromochloromethane	5.01	49	836	0.295	ug/l #	16
29) Tetrahydrofuran	5.12	42	473	0.260	ug/l #	70
36) 1,1-Dichloropropene	5.65	75	28164	4.760	ug/l #	52
38) Carbon Tetrachloride	5.65	117	36674	6.361	ug/l #	14
44) Trichloroethene	7.21	130	846744	165.002	ug/l	98
49) 1,4-Dioxane	7.73	88	360	3.843	ug/l #	45
76) 1,2,3-Trichloropropane	11.13	75	144899	24.305	ug/l #	40
81) trans-1,4-Dichloro-2-buten	11.13	75	144899	62.215	ug/l #	12


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX022620\
Data File : VX015023.D
Acq On : 26 Feb 2020 16:07
Operator : JC/SP
Sample : L1683-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Feb 27 02:56:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X021020W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 10 12:09:20 2020
Response via : Initial Calibration









#1

Pentafluorobenzene
 Concen: 50.000 ug/l
 RT: 5.65 min Scan# 782
 Delta R.T. -0.00 min
 Lab File: VX015023.D
 Acq: 26 Feb 2020 16:07

Instrument :
 MSVOA_X
 ClientSampleId :




Tot Ion:168 Resp: 416676
 Ion Ratio Lower Upper
 168 100
 99 47.3 38.7 58.1




#2

Dichlorodifluoromethane
 Concen: 0.369 ug/l
 RT: 1.19 min Scan# 50
 Delta R.T. -0.00 min
 Lab File: VX015023.D
 Acq: 26 Feb 2020 16:07

Tgt Ion: 85 Resp: 1458
 Ion Ratio Lower Upper
 85 100
 87 42.1 16.1 48.2





#3

Chloromethane

Concen: 0.356 ug/l

RT: 1.32 min Scan# 71

Delta R.T. -0.00 min

Lab File: VX015023.D

Acq: 26 Feb 2020 16:07

Instrument :

MSVOA_X





ClientSampleId :

Tot Ion: 50 Resp: 1591

Ion Ratio Lower Upper

50 100

52 7.0 26.2 39.2#



#4

Vinyl Chloride

Concen: 0.357 ug/l

RT: 1.40 min Scan# 84

Delta R.T. -0.01 min

Lab File: VX015023.D




Acq: 26 Feb 2020 16:07


Tgt Ion: 62 Resp: 1721

Ion Ratio Lower Upper

62 100

64 27.4 24.8 37.2





#5

Bromomethane

Concen: 0.249 ug/l

RT: 1.65 min Scan# 126

Delta R.T. 0.02 min

Lab File: VX015023.D

Acq: 26 Feb 2020 16:07





Instrument :
MSVOA_X
ClientSampled :

Tot Ion: 94 Resp: 776

Ion Ratio Lower Upper

94 100

96 85.5 73.7 110.5



#9

1,1,2-Trichlorotrifluoroethane

Concen: 0.444 ug/l

RT: 2.39 min Scan# 246

Delta R.T. 0.01 min

Lab File: VX015023.D

Acq: 26 Feb 2020 16:07




Tgt Ion: 101 Resp: 1737


Ion Ratio Lower Upper

101 100

85 0.0 34.1 51.1#

151 100.9 67.3 100.9#









#10
Methyl Iodide
Concen: 4.421 ug/l
RT: 2.51 min Scan# 266
Delta R.T. 0.00 min
Lab File: VX015023.D
Acq: 26 Feb 2020 16:07




Instrument :
MSVOA_X
ClientSampled :


Tot Ion:142 Resp: 252
Ion Ratio Lower Upper
142 100
127 175.4 31.1 46.7#
141 81.0 11.3 16.9#



#11
Tert butyl alcohol
Concen: 0.495 ug/l
RT: 3.04 min Scan# 353
Delta R.T. 0.01 min
Lab File: VX015023.D
Acq: 26 Feb 2020 16:07

Tgt Ion: 59 Resp: 411
Ion Ratio Lower Upper
59 100
57 40.1 8.2 12.4#









#12
 1,1-Dichloroethene
 Concen: 1.272 ug/l
 RT: 2.37 min Scan# 243
 Delta R.T. -0.00 min
 Lab File: VX015023.D
 Acq: 26 Feb 2020 16:07




Instrument :
 MSVOA_X
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
96	100		
61	160.4	129.5	194.3
98	57.1	50.6	76.0



#13
 Acrolein
 Concen: 0.613 ug/l
 RT: 2.28 min Scan# 228
 Delta R.T. -0.01 min
 Lab File: VX015023.D
 Acq: 26 Feb 2020 16:07

Tgt Ion	Ratio	Lower	Upper
56	100		
55	0.0	55.0	82.4#









#16
Acetone
Concen: Below Cal
RT: 2.44 min Scan# 254
Delta R.T. 0.01 min
Lab File: VX015023.D
Acq: 26 Feb 2020 16:07




Instrument :
MSVOA_X
ClientSampleId :


Tot Ion: 43 Resp: 3563
Ion Ratio Lower Upper
43 100
58 28.0 24.1 36.1



#21
trans-1,2-Dichloroethene
Concen: 1.444 ug/l
RT: 3.16 min Scan# 373
Delta R.T. -0.00 min
Lab File: VX015023.D
Acq: 26 Feb 2020 16:07

Tgt Ion: 96 Resp: 6084
Ion Ratio Lower Upper
96 100
61 153.7 106.9 160.3
98 70.6 50.7 76.1









#25
2-Butanone
Concen: 0.223 ug/l
RT: 4.65 min Scan# 617
Delta R.T. -0.01 min
Lab File: VX015023.D
Acq: 26 Feb 2020 16:07




Instrument :
MSVOA_X
ClientSampled :


Tot Ion: 43 Resp: 714
Ion Ratio Lower Upper
43 100
72 8.2 21.0 31.6#



#27
cis-1,2-Dichloroethene
Concen: 37.696 ug/l
RT: 4.59 min Scan# 607
Delta R.T. -0.00 min
Lab File: VX015023.D
Acq: 26 Feb 2020 16:07

Tgt Ion: 96 Resp: 177581
Ion Ratio Lower Upper
96 100
61 123.9 0.0 286.2
98 63.7 0.0 131.0









#28
Bromochloromethane
Concen: 0.295 ug/l
RT: 5.01 min Scan# 677
Delta R.T. -0.00 min
Lab File: VX015023.D
Acq: 26 Feb 2020 16:07




Instrument :
MSVOA_X
ClientSampleId :


Tot Ion: 49 Resp: 836
Ion Ratio Lower Upper
49 100
129 29.9 0.0 5.2#
130 7.5 67.4 101.2#



#29
Tetrahydrofuran
Concen: 0.260 ug/l
RT: 5.12 min Scan# 694
Delta R.T. 0.01 min
Lab File: VX015023.D
Acq: 26 Feb 2020 16:07

Tgt Ion: 42 Resp: 473
Ion Ratio Lower Upper
42 100
72 57.7 36.9 55.3#
71 14.6 34.1 51.1#





#33

1,2-Dichloroethane-d4

Concen: 50.795 ug/l

RT: 6.06 min Scan# 848

Delta R.T. -0.00 min

Lab File: VX015023.D

Acq: 26 Feb 2020 16:07

Instrument :

MSVOA_X


ClientSampled :

Tot Ion: 65 Resp: 238866

Ion Ratio Lower Upper




65 100

67 52.1 0.0 103.0



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.86 min Scan# 979

Delta R.T. 0.01 min

Lab File: VX015023.D

Acq: 26 Feb 2020 16:07


Tgt Ion: 114 Resp: 661407

Ion Ratio Lower Upper

114 100

63 19.6 0.0 39.2



88 14.9 0.0 32.0




Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0









#35
 Dibromofluoromethane
 Concen: 50.290 ug/l
 RT: 5.49 min Scan# 755
 Delta R.T. -0.00 min
 Lab File: VX015023.D
 Acq: 26 Feb 2020 16:07




Instrument :
 MSVOA_X
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
113	100		
111	102.6	82.4	123.6
192	24.0	19.1	28.7



#36
 1,1-Dichloropropene
 Concen: 4.760 ug/l
 RT: 5.65 min Scan# 782
 Delta R.T. -0.14 min
 Lab File: VX015023.D
 Acq: 26 Feb 2020 16:07

Tgt Ion	Ratio	Lower	Upper
75	100		
110	11.6	18.3	54.8#
77	1.0	24.6	36.8#





#38

Carbon Tetrachloride

Concen: 6.361 ug/l

RT: 5.65 min Scan# 782

Delta R.T. -0.12 min

Lab File: VX015023.D

Acq: 26 Feb 2020 16:07

Instrument :

MSVOA_X

ClientSampleId :


Tot Ion:117 Resp: 36674

Ion Ratio Lower Upper

117 100

119 2.2 76.0 114.0#




121 0.6 25.4 38.2#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#44

Trichloroethene

Concen: 165.002 ug/l

RT: 7.21 min Scan# 1037

Delta R.T. -0.00 min

Lab File: VX015023.D


Acq: 26 Feb 2020 16:07

Tgt Ion:130 Resp: 846744

Ion Ratio Lower Upper



130 100


95 93.4 0.0 182.6



Abundance Ion 129.80 (129.50 to 130.50): V

Ion 94.90 (94.60 to 95.60): VX0









#49
1,4-Dioxane
Concen: 3.843 ug/l
RT: 7.73 min Scan# 1123
Delta R.T. -0.00 min
Lab File: VX015023.D
Acq: 26 Feb 2020 16:07




Instrument :
MSVOA_X
ClientSampled :


Tot Ion: 88 Resp: 360
Ion Ratio Lower Upper
88 100
43 96.9 27.0 40.4#
58 87.8 53.2 79.8#



#50
Toluene-d8
Concen: 51.373 ug/l
RT: 8.71 min Scan# 1283
Delta R.T. -0.00 min
Lab File: VX015023.D
Acq: 26 Feb 2020 16:07

Tgt Ion: 98 Resp: 797358
Ion Ratio Lower Upper
98 100
100 66.4 52.4 78.6









#62
4-Bromofluorobenzene
Concen: 52.558 ug/l
RT: 11.13 min Scan# 1681
Delta R.T. 0.01 min
Lab File: VX015023.D
Acq: 26 Feb 2020 16:07




Instrument :
MSVOA_X
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
95	100		
174	91.3	0.0	183.0
176	89.1	0.0	178.0



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1513
Delta R.T. -0.00 min
Lab File: VX015023.D
Acq: 26 Feb 2020 16:07

Tgt Ion	Ratio	Lower	Upper
117	100		
82	52.7	40.5	60.7
119	31.8	25.8	38.6









#72

1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.07 min Scan# 1835
Delta R.T. -0.00 min
Lab File: VX015023.D
Acq: 26 Feb 2020 16:07

Instrument :
MSVOA_X
ClientSampled :




Tot Ion	Ratio	Lower	Upper
152	100		
115	54.3	38.6	116.0
150	155.7	0.0	349.0




#76

1,2,3-Trichloropropane
Concen: 24.305 ug/l
RT: 11.13 min Scan# 1681
Delta R.T. -0.16 min
Lab File: VX015023.D
Acq: 26 Feb 2020 16:07

Tgt Ion	Ratio	Lower	Upper
75	100		
77	1.2	18.4	55.0





#81

trans-1,4-Dichloro-2-butene

Concen: 62.215 ug/l

RT: 11.13 min Scan# 1681

Delta R.T. 0.07 min

Lab File: VX015023.D

Acq: 26 Feb 2020 16:07

Instrument :

MSVOA_X

ClientSampleId :


Tot Ion: 75 Resp: 144899

Ion Ratio Lower Upper

75 100

53 0.0 76.2 114.2#

89 0.1 35.8 53.6#



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 53.00 (52.70 to 53.70): VX0

Ion 88.90 (88.60 to 89.60): VX0

